

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/27/2018 15:29

Test: ~~Total~~ ^{Reactive} CN ^{Agb}

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.216	0.0	0.092	
ICB1	-0.113	0.0	0.005	
CCV1	241.461	0.0	0.227	
CCB1	-1.020	0.0	0.004	
PB110640BL	-1.576	0.0	0.003	
J3048-07	-1.536	0.0	0.003	
J3048-11	-1.620	0.0	0.003	
J3048-16	-1.765	0.0	0.003	
J3242-02	-1.409	0.0	0.003	
J3242-04	-1.196	0.0	0.004	
J3242-06	-1.927	0.0	0.003	
J3242-08	-1.757	0.0	0.003	
J3242-10	-1.161	0.0	0.004	
J3242-12	-1.166	0.0	0.004	
CCV2	247.817	0.0	0.233	
CCB2	-0.842	0.0	0.004	
J3242-14	-0.938	0.0	0.004	
J3242-16	-1.008	0.0	0.004	
J3247-08	-0.437	0.0	0.004	
J3247-08DUP	-0.534	0.0	0.004	
J3715-06	-0.786	0.0	0.004	
J3715-12	-0.737	0.0	0.004	
J3734-04	-1.064	0.0	0.004	
CCV3	264.136	0.0	0.248	
CCB3	-0.433	0.0	0.004	
PB110645BL	-1.233	0.0	0.003	
J3048-21	-1.339	0.0	0.003	
J3048-21DUP	-1.463	0.0	0.003	
J3048-23	-1.372	0.0	0.003	
J3717-02	-1.190	0.0	0.004	
J3717-04	-1.301	0.0	0.003	
J3717-06	-1.272	0.0	0.003	
J3717-08	-0.851	0.0	0.004	
CCV4	251.713	0.0	0.236	
CCB4	-0.632	0.0	0.004	

N	35
Mean	30.476
SD	82.1486
CV%	269.55

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/27/2018 12:52


Test Total CN

Accepted 6/27/2018 12:51

Factor 1086
Bias 0.005

Coeff. of det. 0.999801

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.3580	0.0000	
2	5.0PPBCN	0.008	3.7819	5.0000	
3	10PPBCN	0.013	8.8190	10.0000	
4	50PPBCN	0.050	49.4048	50.0000	
5	100PPBCN	0.102	105.8842	100.0000	
6	250PPBCN	0.234	249.0992	250.0000	
7	500PPBCN	0.464	499.3689	500.0000	

Aquakem v. 7.2AQ1

AS

Results from time period:

Wed Jun 27 14:55:23 2018

Wed Jun 27 15:25:52 2018

Sample Id	Sam/Ctr/c	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-1.358	µg/l	6/27/2018 12:49:37	
5.OPPBCN	A	Reactive CN	P	3.7819	µg/l	6/27/2018 12:49:38	
10PPBCN	A	Reactive CN	P	8.819	µg/l	6/27/2018 12:49:39	
50PPBCN	A	Reactive CN	P	49.4048	µg/l	6/27/2018 12:49:40	
100PPBCN	A	Reactive CN	P	105.8842	µg/l	6/27/2018 12:49:41	
250PPBCN	A	Reactive CN	P	249.0992	µg/l	6/27/2018 12:49:42	
500PPBCN	A	Reactive CN	P	499.3689	µg/l	6/27/2018 12:49:43	
ICV1	S	Reactive CN	P	95.2158	µg/l	6/27/2018 14:55:23	
ICB1	S	Reactive CN	P	-0.1132	µg/l	6/27/2018 14:55:24	
CCV1	S	Reactive CN	P	241.4613	µg/l	6/27/2018 14:55:25	
CCB1	S	Reactive CN	P	-1.0198	µg/l	6/27/2018 14:55:26	
PB110640BL	S	Reactive CN	P	-1.5759	µg/l	6/27/2018 14:55:27	
J3048-07	S	Reactive CN	P	-1.5363	µg/l	6/27/2018 14:55:28	
J3048-11	S	Reactive CN	P	-1.6202	µg/l	6/27/2018 14:55:29	
J3048-16	S	Reactive CN	P	-1.7649	µg/l	6/27/2018 14:55:30	
J3242-02	S	Reactive CN	P	-1.4089	µg/l	6/27/2018 14:55:31	
J3242-04	S	Reactive CN	P	-1.1957	µg/l	6/27/2018 14:55:32	
J3242-06	S	Reactive CN	P	-1.9272	µg/l	6/27/2018 14:55:33	
J3242-08	S	Reactive CN	P	-1.7568	µg/l	6/27/2018 15:02:58	
J3242-10	S	Reactive CN	P	-1.1614	µg/l	6/27/2018 15:02:59	
J3242-12	S	Reactive CN	P	-1.1658	µg/l	6/27/2018 15:03:00	
CCV2	S	Reactive CN	P	247.8167	µg/l	6/27/2018 15:03:01	
CCB2	S	Reactive CN	P	-0.8422	µg/l	6/27/2018 15:03:02	
J3242-14	S	Reactive CN	P	-0.9375	µg/l	6/27/2018 15:03:03	
J3242-16	S	Reactive CN	P	-1.0078	µg/l	6/27/2018 15:03:04	
J3247-08	S	Reactive CN	P	-0.4365	µg/l	6/27/2018 15:03:05	
J3247-08DUP	S	Reactive CN	P	-0.5344	µg/l	6/27/2018 15:03:06	
J3715-06	S	Reactive CN	P	-0.7863	µg/l	6/27/2018 15:03:07	
J3715-12	S	Reactive CN	P	-0.7367	µg/l	6/27/2018 15:03:08	
J3734-04	S	Reactive CN	P	-1.0644	µg/l	6/27/2018 15:06:49	
CCV3	S	Reactive CN	P	264.136	µg/l	6/27/2018 15:06:50	
CCB3	S	Reactive CN	P	-0.4333	µg/l	6/27/2018 15:06:51	
PB110645BL	S	Reactive CN	P	-1.2325	µg/l	6/27/2018 15:25:43	
J3048-21	S	Reactive CN	P	-1.3395	µg/l	6/27/2018 15:25:44	
J3048-21DUP	S	Reactive CN	P	-1.4625	µg/l	6/27/2018 15:25:45	
J3048-23	S	Reactive CN	P	-1.3722	µg/l	6/27/2018 15:25:46	
J3717-02	S	Reactive CN	P	-1.1899	µg/l	6/27/2018 15:25:47	
J3717-04	S	Reactive CN	P	-1.3014	µg/l	6/27/2018 15:25:48	
J3717-06	S	Reactive CN	P	-1.272	µg/l	6/27/2018 15:25:49	
J3717-08	S	Reactive CN	P	-0.8511	µg/l	6/27/2018 15:25:50	
CCV4	S	Reactive CN	P	251.7126	µg/l	6/27/2018 15:25:51	
CCB4	S	Reactive CN	P	-0.6323	µg/l	6/27/2018 15:25:52	