

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/28/2018 15:14

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.012	0.0	0.109	
ICB1	-0.001	0.0	0.004	
CCV1	1.016	0.0	0.110	
CCB1	0.002	0.0	0.004	
PB110689BL	0.001	0.0	0.004	
PB110689BS	1.019	0.0	0.110	
J3743-01	0.002	0.0	0.004	
J3743-01DUP	0.001	0.0	0.004	
J3743-01MS	0.955	0.0	0.103	
J3743-01MSD	0.949	0.0	0.103	
CCV2	1.003	0.0	0.108	
CCB2	-0.000	0.0	0.004	

N	12
Mean	0.497
SD	0.5181
CV%	104.32

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Jun 28 14:36:56 2018

Thu Jun 28 15:05:13 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0508	mg/l	6/28/2018 14:36:56	
0.1PPM	A	Phenolics-	P	0.1007	mg/l	6/28/2018 14:36:57	
0.25PPM	A	Phenolics-	P	0.2502	mg/l	6/28/2018 14:36:58	
0.50PPM	A	Phenolics-	P	0.5	mg/l	6/28/2018 14:36:59	
1.0PPM	A	Phenolics-	P	0.9967	mg/l	6/28/2018 14:37:00	
2.0PPM	A	Phenolics-	P	2.0016	mg/l	6/28/2018 14:37:01	
ICV1	S	Phenolics-	P	1.0123	mg/l	6/28/2018 15:00:45	
ICB1	S	Phenolics-	P	-0.0006	mg/l	6/28/2018 15:00:46	
CCV1	S	Phenolics-	P	1.0161	mg/l	6/28/2018 15:00:47	
CCB1	S	Phenolics-	P	0.0024	mg/l	6/28/2018 15:00:48	
PB110689BL	S	Phenolics-	P	0.0011	mg/l	6/28/2018 15:00:49	
PB110689BS	S	Phenolics-	P	1.0192	mg/l	6/28/2018 15:00:50	
J3743-01	S	Phenolics-	P	0.0022	mg/l	6/28/2018 15:00:51	
J3743-01DUP	S	Phenolics-	P	0.0011	mg/l	6/28/2018 15:00:52	
J3743-01MS	S	Phenolics-	P	0.9546	mg/l	6/28/2018 15:00:53	
J3743-01MSD	S	Phenolics-	P	0.9487	mg/l	6/28/2018 15:00:54	
CCV2	S	Phenolics-	P	1.0028	mg/l	6/28/2018 15:00:55	
CCB2	S	Phenolics-	P	-0.0001	mg/l	6/28/2018 15:00:56	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/28/2018 14:37


Test Phenolics-

Accepted 6/28/2018 14:37

Factor 9.581
Bias 0.004

Coeff. of det. 0.999995

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0508	0.0500	
2	Phenol-2	0.014	0.1007	0.1000	
3	Phenol-2	0.030	0.2502	0.2500	
4	Phenol-2	0.056	0.5000	0.5000	
5	Phenol-2	0.108	0.9967	1.0000	
6	Phenol-2	0.213	2.0016	2.0000	