


7470A

L396355
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39603



A= 0.0000e+000

B= 2.5296e-004

C= -1.5503e-002

Rho= 0.9999982

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	61	0.000	61				
0.2	0.200	0.211	0.011	894	0.0 %	894				
2.5	2.500	2.487	-0.013	9894	0.0 %	9894				
5.0	5.000	4.997	-0.003	19816	0.0 %	19816				
7.5	7.500	7.502	0.002	29719	0.0 %	29719				
10.0	10.000	10.003	0.003	39603	0.0 %	39603				

LB96355 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Jun 2018 11:24:20

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	61	-	0.0000	7470A	PPB	29 Jun 2018 11:28:49	S	Std
0.2 - 1	50.2	894	-	0.2000	7470A	PPB	29 Jun 2018 11:30:57	S	Std
2.5 - 1	52.5	9894	-	2.5000	7470A	PPB	29 Jun 2018 11:33:06	S	Std
5.0 - 1	55	19816	-	5.0000	7470A	PPB	29 Jun 2018 11:35:14	S	Std
7.5 - 1	57.5	29719	-	7.5000	7470A	PPB	29 Jun 2018 11:38:56	S	Std
10.0 - 1	510	39603	-	10.0000	7470A	PPB	29 Jun 2018 11:43:26	S	Std
ICV28 - 1	ICV28	15091	3.8020	-	7470A	PPB	29 Jun 2018 11:46:09	U	SMPL
ICB28 - 1	ICB28	-41	-0.0259	-	7470A	PPB	29 Jun 2018 11:48:15	U	SMPL
CCV80 - 1	CCV80	19847	5.0051	-	7470A	PPB	29 Jun 2018 11:50:22	U	SMPL
CCB80 - 1	CCB80	-53	-0.0289	-	7470A	PPB	29 Jun 2018 11:52:29	U	SMPL
CRA - 1	CRA	853	0.2003	-	7470A	PPB	29 Jun 2018 11:54:36	U	SMPL
HighStd - 1	HighStd	39783	10.0481	-	7470A	PPB	29 Jun 2018 11:56:43	U	SMPL
ChkStd - 1	ChkStd	27345	6.9018	-	7470A	PPB	29 Jun 2018 11:58:50	U	SMPL
PB110696BL - 1	PBW	-209	-0.0684	-	7470A	PPB	29 Jun 2018 12:00:59	U	SMPL
PB110696BS - 1	LCSW	13890	3.4982	-	7470A	PPB	29 Jun 2018 12:03:07	U	SMPL
PB110696TB - 1	PB110666TB	-132	-0.0489	-	7470A	PPB	29 Jun 2018 12:05:15	U	SMPL
J3742-01 - 1	JCSA-302	-10	-0.0180	-	7470A	PPB	29 Jun 2018 12:07:23	U	SMPL
J3742-01DUP - 1	JCSA-302DUP	22	-0.0099	-	7470A	PPB	29 Jun 2018 12:09:31	U	SMPL
J3742-01MS - 1	JCSA-302MS	14124	3.5573	-	7470A	PPB	29 Jun 2018 12:14:45	U	SMPL
J3742-01MSD - 1	JCSA-302MSD	14204	3.5776	-	7470A	PPB	29 Jun 2018 12:16:51	U	SMPL
CCV81 - 1	CCV81	19827	5.0000	-	7470A	PPB	29 Jun 2018 12:18:57	U	SMPL
CCB81 - 1	CCB81	-73	-0.0340	-	7470A	PPB	29 Jun 2018 12:21:04	U	SMPL
PB110697BL - 1	PBW	-8	-0.0175	-	7470A	PPB	29 Jun 2018 12:23:11	U	SMPL
PB110697BS - 1	LCSW	12841	3.2328	-	7470A	PPB	29 Jun 2018 12:25:18	U	SMPL
J3743-01 - 1	GMW-5	3313	0.8226	-	7470A	PPB	29 Jun 2018 12:27:24	U	SMPL
J3743-01DUP - 1	GMW-5DUP	3003	0.7441	-	7470A	PPB	29 Jun 2018 12:34:06	U	SMPL
J3743-01MS - 1	GMW-5MS	18084	4.5591	-	7470A	PPB	29 Jun 2018 12:36:12	U	SMPL
J3743-01MSD - 1	GMW-5MSD	17584	4.4326	-	7470A	PPB	29 Jun 2018 12:38:20	U	SMPL
J3750-03 - 1	ZER-GW-SB-03	20331	5.1275	-	7470A	PPB	29 Jun 2018 12:40:29	U	SMPL
J3755-01 - 1	PS-BASELINEWATER	-358	-0.1061	-	7470A	PPB	29 Jun 2018 12:42:38	U	SMPL
CCV82 - 1	CCV82	19770	4.9856	-	7470A	PPB	29 Jun 2018 12:44:47	U	SMPL
CCB82 - 1	CCB82	-25	-0.0218	-	7470A	PPB	29 Jun 2018 12:46:54	U	SMPL
J3742-01LX5 - 1		37	-0.0061	-	7470A	PPB	29 Jun 2018 12:49:02	U	SMPL
J3742-01A - 1		14879	3.7483	-	7470A	PPB	29 Jun 2018 12:51:09	U	SMPL
J3743-01LX5 - 1		457	0.1001	-	7470A	PPB	29 Jun 2018 12:53:16	U	SMPL
J3743-01A - 1		19058	4.8055	-	7470A	PPB	29 Jun 2018 12:55:23	U	SMPL
CCV83 - 1	CCV83	19492	4.9153	-	7470A	PPB	29 Jun 2018 12:57:29	U	SMPL
CCB83 - 1	CCB83	-70	-0.0332	-	7470A	PPB	29 Jun 2018 12:59:37	U	SMPL