

2B96511

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/9/2018 14:38

Test: *Reactive* ~~Total~~ CN *µg/L*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.964	0.0	0.091	
ICB1	0.113	0.0	0.004	
CCV1	238.479	0.0	0.217	
CCB1	-0.139	0.0	0.003	
PB110936BL	-0.135	0.0	0.003	
J3891-04	-0.495	0.0	0.003	
J3891-04DUP	0.039	0.0	0.004	
J3891-08	-0.451	0.0	0.003	
J3893-04	-0.269	0.0	0.003	
J3896-04	-1.144	0.0	0.003	
J3897-06	-0.322	0.0	0.003	
J3901-04	-0.561	0.0	0.003	
CCV2	257.620	0.0	0.234	
CCB2	-0.041	0.0	0.004	

N 14
Mean 42.190
SD 91.1126
CV% 215.96

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 09 14:34:24 2018

Mon Jul 09 14:38:17 2018

Sample Id	Sam/Ctr/c	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.8285	µg/l	7/9/2018 9:58:00	
5.0PPBCN	A	Reactive CN	P	5.4414	µg/l	7/9/2018 9:58:01	
10PPBCN	A	Reactive CN	P	9.5099	µg/l	7/9/2018 9:58:02	
50PPBCN	A	Reactive CN	P	50.5026	µg/l	7/9/2018 9:58:03	
100PPBCN	A	Reactive CN	P	101.5287	µg/l	7/9/2018 9:58:04	
250PPBCN	A	Reactive CN	P	248.3929	µg/l	7/9/2018 9:58:05	
500PPBCN	A	Reactive CN	P	500.4529	µg/l	7/9/2018 9:58:06	
ICV1	S	Reactive CN	P	97.9635	µg/l	7/9/2018 14:34:24	
ICB1	S	Reactive CN	P	0.1126	µg/l	7/9/2018 14:34:25	
CCV1	S	Reactive CN	P	238.4788	µg/l	7/9/2018 14:34:26	
CCB1	S	Reactive CN	P	-0.1388	µg/l	7/9/2018 14:34:27	
PB110936BL	S	Reactive CN	P	-0.1352	µg/l	7/9/2018 14:34:28	
J3891-04	S	Reactive CN	P	-0.4946	µg/l	7/9/2018 14:34:29	
J3891-04DUP	S	Reactive CN	P	0.0387	µg/l	7/9/2018 14:34:30	
J3891-08	S	Reactive CN	P	-0.4506	µg/l	7/9/2018 14:34:31	
J3893-04	S	Reactive CN	P	-0.269	µg/l	7/9/2018 14:34:32	
J3896-04	S	Reactive CN	P	-1.1437	µg/l	7/9/2018 14:34:33	
J3897-06	S	Reactive CN	P	-0.3223	µg/l	7/9/2018 14:34:34	
J3901-04	S	Reactive CN	P	-0.5612	µg/l	7/9/2018 14:38:15	
CCV2	S	Reactive CN	P	257.6198	µg/l	7/9/2018 14:38:16	
CCB2	S	Reactive CN	P	-0.0413	µg/l	7/9/2018 14:38:17	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/9/2018 10:04


Test Total CN

Accepted 7/9/2018 10:04

Factor 1118
Bias 0.004

Coeff. of det. 0.999968

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.8285	0.0000	
2	5.0PPBCN	0.008	5.4414	5.0000	
3	10PPBCN	0.012	9.5099	10.0000	
4	50PPBCN	0.049	50.5026	50.0000	
5	100PPBCN	0.094	101.5287	100.0000	
6	250PPBCN	0.226	248.3929	250.0000	
7	500PPBCN	0.451	500.4529	500.0000	