

LB96525

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/10/2018 9:02

Test: Phenolics mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.022	0.0	0.105	
ICB1	-0.002	0.0	0.003	
CCV1	0.996	0.0	0.103	
CCB1	-0.000	0.0	0.004	
PB110941BL	0.001	0.0	0.004	
PB110941BS	1.024	0.0	0.105	
J3762-10	0.009	0.0	0.004	
J3762-11	0.090	0.0	0.012	
J3762-12	0.010	0.0	0.005	
CCV2	0.977	0.0	0.101	
CCB2	-0.001	0.0	0.003	

N 11
Mean 0.375
SD 0.5001
CV% 133.30

48

Aquakem v. 7.2AQ1

Results from time period:

Tue Jul 10 08:53:51 2018

Tue Jul 10 08:57:50 2018

Sample Id	Sam/Ctr/c	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0497	mg/l	7/10/2018 7:45:00	
0.1PPM	A	Phenolics-	P	0.1	mg/l	7/10/2018 7:45:01	
0.25PPM	A	Phenolics-	P	0.2497	mg/l	7/10/2018 7:45:02	
0.50PPM	A	Phenolics-	P	0.4976	mg/l	7/10/2018 7:45:03	
1.0PPM	A	Phenolics-	P	1.0048	mg/l	7/10/2018 7:45:04	
2.0PPM	A	Phenolics-	P	1.9983	mg/l	7/10/2018 7:45:05	
ICV1	S	Phenolics-	P	1.0222	mg/l	7/10/2018 8:53:51	
ICB1	S	Phenolics-	P	-0.0016	mg/l	7/10/2018 8:53:52	
CCV1	S	Phenolics-	P	0.9957	mg/l	7/10/2018 8:53:53	
CCB1	S	Phenolics-	P	0	mg/l	7/10/2018 8:53:54	
PB110941BL	S	Phenolics-	P	0.0014	mg/l	7/10/2018 8:53:55	
PB110941BS	S	Phenolics-	P	1.0242	mg/l	7/10/2018 8:53:56	
J3762-10	S	Phenolics-	P	0.0087	mg/l	7/10/2018 8:53:58	
J3762-11	S	Phenolics-	P	0.0901	mg/l	7/10/2018 8:53:59	
J3762-12	S	Phenolics-	P	0.01	mg/l	7/10/2018 8:54:00	
CCV2	S	Phenolics-	P	0.9774	mg/l	7/10/2018 8:54:01	
CCB2	S	Phenolics-	P	-0.0015	mg/l	7/10/2018 8:54:02	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/10/2018 7:45


Test Phenolics-

Accepted 7/10/2018 7:45

Factor 10.06
Bias 0.004

Coeff. of det. 0.999989

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0497	0.0500	
2	Phenol-2	0.013	0.1000	0.1000	
3	Phenol-2	0.028	0.2497	0.2500	
4	Phenol-2	0.053	0.4976	0.5000	
5	Phenol-2	0.103	1.0048	1.0000	
6	Phenol-2	0.202	1.9983	2.0000	