

7470A


LB96529

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38533



A= 0.0000e+000

B= 2.5971e-004

C= -4.4420e-002

Rho= 0.9999081

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.063	-0.063	-71	0.000	-71				
0.2	0.200	0.204	0.004	958	0.0 %	958				
2.5	2.500	2.526	0.026	9899	0.0 %	9899				
5.0	5.000	5.092	0.092	19777	0.0 %	19777				
7.5	7.500	7.477	-0.023	28960	0.0 %	28960				
10.0	10.000	9.963	-0.037	38533	0.0 %	38533				

LB96529 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Jul 2018 11:56:03

Sample ID	Extended ID	μ Abs	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-71	-	0.00007470A	PPB	10 Jul 2018 12:00:00	S	Std
0.2 - 1	50.2	958	-	0.20007470A	PPB	10 Jul 2018 12:06:44	S	Std
2.5 - 1	52.5	9899	-	2.50007470A	PPB	10 Jul 2018 12:08:52	S	Std
5.0 - 1	55	19777	-	5.00007470A	PPB	10 Jul 2018 12:11:00	S	Std
7.5 - 1	57.5	28960	-	7.50007470A	PPB	10 Jul 2018 12:13:08	S	Std
10.0 - 1	510	38533	-	10.00007470A	PPB	10 Jul 2018 12:15:16	S	Std
ICV41 - 1	ICV41	14411	3.6983	-7470A	PPB	10 Jul 2018 12:17:52	U	SMPL
ICB41 - 1	ICB41	-83	-0.0660	-7470A	PPB	10 Jul 2018 12:19:58	U	SMPL
CCV40 - 1	CCV40	19795	5.0966	-7470A	PPB	10 Jul 2018 12:22:04	U	SMPL
CCB40 - 1	CCB40	-88	-0.0673	-7470A	PPB	10 Jul 2018 12:24:11	U	SMPL
CRA - 1	CRA	976	0.2091	-7470A	PPB	10 Jul 2018 12:26:17	U	SMPL
HighStd - 1	HighStd	38210	9.8793	-7470A	PPB	10 Jul 2018 12:28:24	U	SMPL
ChkStd - 1	ChkStd	26996	6.9668	-7470A	PPB	10 Jul 2018 12:30:31	U	SMPL
PB110914BL - 1	PBW	-152	-0.0839	-7470A	PPB	10 Jul 2018 12:32:40	U	SMPL
PB110914BS - 1	LCSW	15084	3.8731	-7470A	PPB	10 Jul 2018 12:37:22	U	SMPL
PB110914TB - 1	PB110864TB	83	-0.0229	-7470A	PPB	10 Jul 2018 12:39:30	U	SMPL
J3874-03 - 1	2018-NYSEG-MCMaster-PH4-5(0-3)	102	-0.0179	-7470A	PPB	10 Jul 2018 12:41:38	U	SMPL
J3874-04 - 1	2018-NYSEG-MCMaster-PH4-5(3-5)	-98	-0.0699	-7470A	PPB	10 Jul 2018 12:43:46	U	SMPL
J3877-02 - 1	NS-DRUM-2-STONE	176	0.0013	-7470A	PPB	10 Jul 2018 12:45:55	U	SMPL
J3887-04 - 1	TP-8-S	52	-0.0309	-7470A	PPB	10 Jul 2018 12:48:02	U	SMPL
CCV41 - 1	CCV41	19916	5.1280	-7470A	PPB	10 Jul 2018 12:50:08	U	SMPL
CCB41 - 1	CCB41	-93	-0.0686	-7470A	PPB	10 Jul 2018 12:52:15	U	SMPL
J3887-08 - 1	TP-8-D	46	-0.0325	-7470A	PPB	10 Jul 2018 12:54:22	U	SMPL
J3887-08DUP - 1	TP-8-DDUP	74	-0.0252	-7470A	PPB	10 Jul 2018 12:56:29	U	SMPL
J3887-08MS - 1	TP-8-DMS	16125	4.1435	-7470A	PPB	10 Jul 2018 12:58:36	U	SMPL
J3887-08MSD - 1	TP-8-DMSD	16072	4.1297	-7470A	PPB	10 Jul 2018 13:00:43	U	SMPL
PB110933BL - 1	PBW	-99	-0.0701	-7470A	PPB	10 Jul 2018 13:05:20	U	SMPL
PB110933BS - 1	LCSW	13329	3.4173	-7470A	PPB	10 Jul 2018 13:07:27	U	SMPL
PB110933TB - 1	PB110910TB	25	-0.0379	-7470A	PPB	10 Jul 2018 13:12:03	U	SMPL
J3891-04 - 1	212-72-11940-R4-VNJ-242	-144	-0.0818	-7470A	PPB	10 Jul 2018 13:14:11	U	SMPL
J3891-04DUP - 1	212-72-11940-R4-VNJ-242DUP	-150	-0.0834	-7470A	PPB	10 Jul 2018 13:16:19	U	SMPL
J3891-04MS - 1	212-72-11940-R4-VNJ-242MS	13111	3.3607	-7470A	PPB	10 Jul 2018 13:18:26	U	SMPL
CCV42 - 1	CCV42	19600	5.0460	-7470A	PPB	10 Jul 2018 13:20:33	U	SMPL
CCB42 - 1	CCB42	-92	-0.0683	-7470A	PPB	10 Jul 2018 13:22:40	U	SMPL
J3891-04MSD - 1	212-72-11940-R4-VNJ-242MSD	13121	3.3633	-7470A	PPB	10 Jul 2018 13:24:48	U	SMPL
J3891-08 - 1	72-12002-RBR-RBR-VNJ-230	-235	-0.1055	-7470A	PPB	10 Jul 2018 13:26:55	U	SMPL
J3893-04 - 1	BS-S-7-SOIL	-155	-0.0847	-7470A	PPB	10 Jul 2018 13:29:02	U	SMPL
J3896-04 - 1	JC-SOIL	-145	-0.0821	-7470A	PPB	10 Jul 2018 13:31:10	U	SMPL
J3897-06 - 1	STOCK-PILE-A	-139	-0.0805	-7470A	PPB	10 Jul 2018 13:33:17	U	SMPL
J3901-04 - 1	TP-9-S	-138	-0.0803	-7470A	PPB	10 Jul 2018 13:35:25	U	SMPL
J3887-08LX5 - 1		-31	-0.0525	-7470A	PPB	10 Jul 2018 13:37:33	U	SMPL
J3887-08A - 1		14227	3.6505	-7470A	PPB	10 Jul 2018 13:39:41	U	SMPL
J3891-04LX5 - 1		-172	-0.0891	-7470A	PPB	10 Jul 2018 13:41:50	U	SMPL
J3891-04A - 1		14257	3.6583	-7470A	PPB	10 Jul 2018 13:43:56	U	SMPL
CCV43 - 1	CCV43	19729	5.0795	-7470A	PPB	10 Jul 2018 13:46:04	U	SMPL
CCB43 - 1	CCB43	-83	-0.0660	-7470A	PPB	10 Jul 2018 13:48:11	U	SMPL