

LB 96535

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/10/2018 15:00

Test: Total CN *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.074	0.0	0.094	
ICB1	1.981	0.0	0.006	
CCV1	242.245	0.0	0.228	
CCB1	-0.190	0.0	0.004	
PB110973BL	-0.008	0.0	0.004	
PB110973BS	98.199	0.0	0.095	
J3911-01MS	31.988	0.0	0.034	
J3911-01MSD	31.608	0.0	0.033	
J3911-01	8.316	0.0	0.012	
J3911-01DUP	8.595	0.0	0.012	
J3911-03	5.397	0.0	0.009	
CCV2	262.019	0.0	0.246	
CCB2	-0.372	0.0	0.004	

N	13
Mean	60.527
SD	91.7313
CV%	151.55

Aquakem v. 7.2AQ1

AS

Results from time period:

Tue Jul 10 14:53:03 2018

Tue Jul 10 14:56:27 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.9166	µg/l	7/10/2018 10:27:03	
5.0PPBCN	A	Total CN	P	4.2871	µg/l	7/10/2018 10:27:04	
10PPBCN	A	Total CN	P	9.5997	µg/l	7/10/2018 10:27:05	
50PPBCN	A	Total CN	P	49.3272	µg/l	7/10/2018 10:27:06	
100PPBCN	A	Total CN	P	101.4044	µg/l	7/10/2018 10:27:07	
250PPBCN	A	Total CN	P	252.9933	µg/l	7/10/2018 10:27:08	
500PPBCN	A	Total CN	P	498.3049	µg/l	7/10/2018 10:27:09	
ICV1	S	Total CN	P	97.0742	µg/l	7/10/2018 14:53:03	
ICB1	S	Total CN	P	1.9812	µg/l	7/10/2018 14:53:04	
CCV1	S	Total CN	P	242.2445	µg/l	7/10/2018 14:53:05	
CCB1	S	Total CN	P	-0.1898	µg/l	7/10/2018 14:53:06	
PB110973BL	S	Total CN	P	-0.0081	µg/l	7/10/2018 14:53:07	
PB110973BS	S	Total CN	P	98.1986	µg/l	7/10/2018 14:53:08	
J3911-01MS	S	Total CN	P	31.9878	µg/l	7/10/2018 14:53:09	
J3911-01MSD	S	Total CN	P	31.608	µg/l	7/10/2018 14:53:10	
J3911-01	S	Total CN	P	8.3155	µg/l	7/10/2018 14:53:11	
J3911-01DUP	S	Total CN	P	8.5952	µg/l	7/10/2018 14:53:12	
J3911-03	S	Total CN	P	5.3968	µg/l	7/10/2018 14:53:13	
CCV2	S	Total CN	P	262.019	µg/l	7/10/2018 14:56:26	
CCB2	S	Total CN	P	-0.3718	µg/l	7/10/2018 14:56:27	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/10/2018 10:34


Test Total CN

Accepted 7/10/2018 10:34

Factor 1081
Bias 0.004

Coeff. of det. 0.999923

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.9166	0.0000	
2	5.0PPBCN	0.008	4.2871	5.0000	
3	10PPBCN	0.013	9.5997	10.0000	
4	50PPBCN	0.050	49.3272	50.0000	
5	100PPBCN	0.098	101.4044	100.0000	
6	250PPBCN	0.238	252.9933	250.0000	
7	500PPBCN	0.465	498.3049	500.0000	