

=====21396662=====

Test results Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/16/2018 12:27

Test: Reactive ~~Total~~ CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	91.569	0.0	0.089	
ICB1	0.939	0.0	0.004	
CCV1	246.918	0.0	0.234	
CCB1	1.089	0.0	0.004	
PB111145BL	0.681	0.0	0.004	
J3819-02	1.008	0.0	0.004	
J3819-02DUP	0.842	0.0	0.004	
J3819-04	0.416	0.0	0.003	
J3819-06	0.622	0.0	0.004	
J3819-08	0.793	0.0	0.004	
J3829-02	0.561	0.0	0.003	
J3829-04	0.591	0.0	0.003	
J3829-06	0.636	0.0	0.004	
CCV2	245.391	0.0	0.233	
CCB2	0.860	0.0	0.004	
J3997-04	0.646	0.0	0.004	
J3997-08	0.447	0.0	0.003	
J3997-12	0.805	0.0	0.004	
J3997-16	1.031	0.0	0.004	
J3996-06	0.713	0.0	0.004	
CCV3	265.030	0.0	0.251	
CCB3	1.228	0.0	0.004	

N	22
Mean	39.219
SD	88.9045
CV%	226.69

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/16/2018 12:10


Test Total CN

Accepted 7/16/2018 8:47

Factor 1066
Bias 0.003

Coeff. of det. 0.999684

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.4625	0.0000	
2	5.0PPBCN	0.008	5.7971	5.0000	
3	10PPBCN	0.013	10.5948	10.0000	
4	50PPBCN	0.050	50.0738	50.0000	
5	100PPBCN	0.091	94.2785	100.0000	
6	250PPBCN	0.242	255.3526	250.0000	
7	500PPBCN	0.470	498.4408	500.0000	

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 16 11:59:06 2018

Mon Jul 16 12:22:28 2018

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.4625	µg/l	7/16/2018 8:23:29	
5.OPPBCN	A	Reactive CN	P	5.7971	µg/l	7/16/2018 8:23:30	
10PPBCN	A	Reactive CN	P	10.5948	µg/l	7/16/2018 8:23:31	
50PPBCN	A	Reactive CN	P	50.0738	µg/l	7/16/2018 8:23:32	
100PPBCN	A	Reactive CN	P	94.2785	µg/l	7/16/2018 8:23:33	
250PPBCN	A	Reactive CN	P	255.3526	µg/l	7/16/2018 8:23:34	
500PPBCN	A	Reactive CN	P	498.4408	µg/l	7/16/2018 8:23:35	
ICV1	S	Reactive CN	P	91.5694	µg/l	7/16/2018 11:59:06	
ICB1	S	Reactive CN	P	0.9387	µg/l	7/16/2018 11:59:07	
CCV1	S	Reactive CN	P	246.9181	µg/l	7/16/2018 11:59:08	
CCB1	S	Reactive CN	P	1.0885	µg/l	7/16/2018 11:59:09	
PB111145BL	S	Reactive CN	P	0.6808	µg/l	7/16/2018 11:59:10	
J3819-02	S	Reactive CN	P	1.0081	µg/l	7/16/2018 11:59:11	
J3819-02DUP	S	Reactive CN	P	0.8419	µg/l	7/16/2018 11:59:12	
J3819-04	S	Reactive CN	P	0.4155	µg/l	7/16/2018 11:59:13	
J3819-06	S	Reactive CN	P	0.6219	µg/l	7/16/2018 11:59:14	
J3819-08	S	Reactive CN	P	0.7927	µg/l	7/16/2018 11:59:15	
J3829-02	S	Reactive CN	P	0.5612	µg/l	7/16/2018 11:59:16	
J3829-04	S	Reactive CN	P	0.5906	µg/l	7/16/2018 12:06:41	
J3829-06	S	Reactive CN	P	0.6357	µg/l	7/16/2018 12:06:42	
CCV2	S	Reactive CN	P	245.3914	µg/l	7/16/2018 12:06:44	
CCB2	S	Reactive CN	P	0.8598	µg/l	7/16/2018 12:06:45	
J3997-04	S	Reactive CN	P	0.6457	µg/l	7/16/2018 12:06:46	
J3997-08	S	Reactive CN	P	0.4471	µg/l	7/16/2018 12:06:47	
J3997-12	S	Reactive CN	P	0.8049	µg/l	7/16/2018 12:06:48	
J3997-16	S	Reactive CN	P	1.0308	µg/l	7/16/2018 12:06:49	
J3996-06	S	Reactive CN	P	0.7131	µg/l	7/16/2018 12:22:26	
CCV3	S	Reactive CN	P	265.03	µg/l	7/16/2018 12:22:27	
CCB3	S	Reactive CN	P	1.2279	µg/l	7/16/2018 12:22:28	