

21396786

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/20/2018 12:45

Test: *Reactive*
~~Total~~ CN *Mg/L*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.897	0.0	0.095	
ICB1	-0.696	0.0	0.004	
CCV1	247.199	0.0	0.233	
CCB1	-0.315	0.0	0.004	
PB111311BL	-0.954	0.0	0.004	
J3820-06	-1.133	0.0	0.004	
J3820-06DUP	-1.266	0.0	0.004	
J3827-04	-1.016	0.0	0.004	
J4081-03	-0.966	0.0	0.004	
J4087-53	-1.083	0.0	0.004	
CCB2	-1.090	0.0	0.004	
PB111312BL	-1.098	0.0	0.004	
J4066-02	-1.176	0.0	0.004	
J4066-02DUP	-0.917	0.0	0.004	
CCV2	263.543	0.0	0.248	
CCB2	-0.579	0.0	0.004	

N 16
Mean 37.272
SD 88.6685
CV% 237.90

AS
07/20/18

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/20/2018 9:33


Test Total CN

Accepted 7/20/2018 9:33

Factor 1083
Bias 0.005

Coeff. of det. 0.999621

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.5413	0.0000	
2	5.0PPBCN	0.009	4.1865	5.0000	
3	10PPBCN	0.012	8.1638	10.0000	
4	50PPBCN	0.050	49.2567	50.0000	
5	100PPBCN	0.098	101.4178	100.0000	
6	250PPBCN	0.242	257.3616	250.0000	
7	500PPBCN	0.463	496.1548	500.0000	

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Jul 20 11:41:58 2018

Fri Jul 20 12:45:13 2018

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-1.5413	µg/l	7/20/2018 9:32:41	
5.OPPBCN	A	Reactive CN	P	4.1865	µg/l	7/20/2018 9:32:42	
10PPBCN	A	Reactive CN	P	8.1638	µg/l	7/20/2018 9:32:43	
50PPBCN	A	Reactive CN	P	49.2567	µg/l	7/20/2018 9:32:44	
100PPBCN	A	Reactive CN	P	101.4178	µg/l	7/20/2018 9:32:45	
250PPBCN	A	Reactive CN	P	257.3616	µg/l	7/20/2018 9:32:46	
500PPBCN	A	Reactive CN	P	496.1548	µg/l	7/20/2018 9:32:47	
ICV1	S	Reactive CN	P	97.8969	µg/l	7/20/2018 12:07:03	
ICB1	S	Reactive CN	P	-0.696	µg/l	7/20/2018 12:07:04	
CCV1	S	Reactive CN	P	247.1992	µg/l	7/20/2018 12:07:05	
CCB1	S	Reactive CN	P	-0.3155	µg/l	7/20/2018 12:07:06	
PB111311BL	S	Reactive CN	P	-0.9543	µg/l	7/20/2018 12:07:07	
J3820-06	S	Reactive CN	P	-1.133	µg/l	7/20/2018 12:07:08	
J3820-06DUP	S	Reactive CN	P	-1.2663	µg/l	7/20/2018 12:07:09	
J3827-04	S	Reactive CN	P	-1.0158	µg/l	7/20/2018 12:07:10	
J4081-03	S	Reactive CN	P	-0.9658	µg/l	7/20/2018 12:07:11	
J4087-53	S	Reactive CN	P	-1.0828	µg/l	7/20/2018 12:07:12	
PB111312BL	S	Reactive CN	P	-1.0976	µg/l	7/20/2018 12:11:22	
J4066-02	S	Reactive CN	P	-1.176	µg/l	7/20/2018 12:11:23	
J4066-02DUP	S	Reactive CN	P	-0.9171	µg/l	7/20/2018 12:11:24	
CCV2	S	Reactive CN	P	263.543	µg/l	7/20/2018 12:45:12	
CCB2	S	Total CN	P	-0.5788	µg/l	7/20/2018 12:45:13	