

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/23/2018 12:03

Test: ^{Reactive} Total CN ug/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.455	0.0	0.095	
ICB1	0.394	0.0	0.004	
CCV1	241.115	0.0	0.232	
CCB1	0.234	0.0	0.004	
PB111360BL	0.050	0.0	0.003	
J4101-02	0.262	0.0	0.004	
J4101-02DUP	0.578	0.0	0.004	
J4101-04	0.116	0.0	0.004	
J4102-01	0.233	0.0	0.004	
J4102-02	0.093	0.0	0.004	
PB111361BL	-0.005	0.0	0.003	
J4102-03	0.012	0.0	0.003	
J4102-03DUP	0.796	0.0	0.004	
J4102-04	0.526	0.0	0.004	
CCV2	251.683	0.0	0.242	
CCB2	0.761	0.0	0.004	
J4102-05	0.504	0.0	0.004	
CCV3	251.363	0.0	0.242	
CCB3	0.109	0.0	0.004	

N	19
Mean	44.488
SD	93.2037
CV%	209.50

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/23/2018 10:41


Test Total CN

Accepted 7/23/2018 10:41

Factor 1054
Bias 0.003

Coeff. of det. 0.999939

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.2634	0.0000	
2	5.0PPBCN	0.008	4.7847	5.0000	
3	10PPBCN	0.013	10.0872	10.0000	
4	50PPBCN	0.051	50.5992	50.0000	
5	100PPBCN	0.099	101.0642	100.0000	
6	250PPBCN	0.238	246.9475	250.0000	
7	500PPBCN	0.479	501.2540	500.0000	

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 23 11:57:13 2018

Mon Jul 23 12:03:31 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	0.2634	µg/l	7/23/2018 10:41:08	
5.0PPBCN	A	Reactive CN	P	4.7847	µg/l	7/23/2018 10:41:09	
10PPBCN	A	Reactive CN	P	10.0872	µg/l	7/23/2018 10:41:10	
50PPBCN	A	Reactive CN	P	50.5992	µg/l	7/23/2018 10:41:11	
100PPBCN	A	Reactive CN	P	101.0642	µg/l	7/23/2018 10:41:12	
250PPBCN	A	Reactive CN	P	246.9475	µg/l	7/23/2018 10:41:13	
500PPBCN	A	Reactive CN	P	501.254	µg/l	7/23/2018 10:41:14	
ICV1	S	Reactive CN	P	96.4546	µg/l	7/23/2018 11:57:13	
ICB1	S	Reactive CN	P	0.3938	µg/l	7/23/2018 11:57:14	
CCV1	S	Reactive CN	P	241.115	µg/l	7/23/2018 11:57:15	
CCB1	S	Reactive CN	P	0.2339	µg/l	7/23/2018 11:57:16	
PB111360BL	S	Reactive CN	P	0.0505	µg/l	7/23/2018 11:57:17	
J4101-02	S	Reactive CN	P	0.2625	µg/l	7/23/2018 11:57:18	
J4101-02DUP	S	Reactive CN	P	0.5779	µg/l	7/23/2018 11:57:19	
J4101-04	S	Reactive CN	P	0.1163	µg/l	7/23/2018 11:57:20	
J4102-01	S	Reactive CN	P	0.2332	µg/l	7/23/2018 11:57:21	
J4102-02	S	Reactive CN	P	0.0931	µg/l	7/23/2018 11:57:22	
PB111361BL	S	Reactive CN	P	-0.005	µg/l	7/23/2018 11:57:23	
J4102-03	S	Reactive CN	P	0.0123	µg/l	7/23/2018 12:03:24	
J4102-03DUP	S	Reactive CN	P	0.7958	µg/l	7/23/2018 12:03:25	
J4102-04	S	Reactive CN	P	0.5262	µg/l	7/23/2018 12:03:26	
CCV2	S	Reactive CN	P	251.683	µg/l	7/23/2018 12:03:27	
CCB2	S	Reactive CN	P	0.7609	µg/l	7/23/2018 12:03:28	
J4102-05	S	Reactive CN	P	0.5044	µg/l	7/23/2018 12:03:29	
CCV3	S	Reactive CN	P	251.3628	µg/l	7/23/2018 12:03:30	
CCB3	S	Reactive CN	P	0.1091	µg/l	7/23/2018 12:03:31	