

21396859

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/25/2018 14:13

Test: *Reactive*
~~Total CN~~ *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.095	0.0	0.091	
ICB1	1.645	0.0	0.004	
CCV1	236.143	0.0	0.221	
CCB1	1.415	0.0	0.004	
PB111455BL	1.144	0.0	0.004	
J3821-08	1.142	0.0	0.004	
J3821-08DUP	1.180	0.0	0.004	
J3821-12	0.998	0.0	0.003	
J4123-04	1.263	0.0	0.004	
J4124-04	1.161	0.0	0.004	
J4124-08	1.152	0.0	0.004	
J4135-01	0.937	0.0	0.003	
J4135-02	1.387	0.0	0.004	
CCV2	244.412	0.0	0.229	
CCB2	1.418	0.0	0.004	
J4135-03	1.059	0.0	0.003	
J3821-10	1.047	0.0	0.003	
CCV3	249.730	0.0	0.234	
CCB3	1.311	0.0	0.004	
N	19			
Mean	44.402			
SD	91.1264			
CV%	205.23			

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/25/2018 13:11


Test Total CN

Accepted 7/25/2018 13:11

Factor 1079
Bias 0.002

Coeff. of det. 0.999852

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.5314	0.0000	
2	5.0PPBCN	0.008	5.6819	5.0000	
3	10PPBCN	0.013	11.0584	10.0000	
4	50PPBCN	0.050	50.9885	50.0000	
5	100PPBCN	0.094	98.9358	100.0000	
6	250PPBCN	0.230	245.4359	250.0000	
7	500PPBCN	0.468	502.3681	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Wed Jul 25 13:05:47 2018

Wed Jul 25 14:02:11 2018

AS

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.5314	µg/l	7/25/2018 13:05:47	
5.OPPBCN	A	Reactive CN	P	5.6819	µg/l	7/25/2018 13:05:48	
10PPBCN	A	Reactive CN	P	11.0584	µg/l	7/25/2018 13:05:49	
50PPBCN	A	Reactive CN	P	50.9885	µg/l	7/25/2018 13:05:50	
100PPBCN	A	Reactive CN	P	98.9358	µg/l	7/25/2018 13:05:51	
250PPBCN	A	Reactive CN	P	245.4359	µg/l	7/25/2018 13:05:52	
500PPBCN	A	Reactive CN	P	502.3681	µg/l	7/25/2018 13:05:53	
ICV1	S	Reactive CN	P	95.0948	µg/l	7/25/2018 13:39:24	
ICB1	S	Reactive CN	P	1.6447	µg/l	7/25/2018 13:39:25	
CCV1	S	Reactive CN	P	236.1425	µg/l	7/25/2018 13:39:26	
CCB1	S	Reactive CN	P	1.4146	µg/l	7/25/2018 13:39:27	
PB111455BL	S	Reactive CN	P	1.1442	µg/l	7/25/2018 13:39:28	
J3821-08	S	Reactive CN	P	1.1419	µg/l	7/25/2018 13:39:29	
J3821-08DUP	S	Reactive CN	P	1.1798	µg/l	7/25/2018 13:39:30	
J3821-12	S	Reactive CN	P	0.9976	µg/l	7/25/2018 13:39:32	
J4123-04	S	Reactive CN	P	1.2627	µg/l	7/25/2018 13:39:33	
J4124-04	S	Reactive CN	P	1.161	µg/l	7/25/2018 13:39:34	
J4124-08	S	Reactive CN	P	1.1518	µg/l	7/25/2018 13:45:35	
J4135-01	S	Reactive CN	P	0.9371	µg/l	7/25/2018 13:45:36	
J4135-02	S	Reactive CN	P	1.3867	µg/l	7/25/2018 13:45:37	
CCV2	S	Reactive CN	P	244.4123	µg/l	7/25/2018 13:45:38	
CCB2	S	Reactive CN	P	1.4183	µg/l	7/25/2018 13:45:39	
J4135-03	S	Reactive CN	P	1.0593	µg/l	7/25/2018 13:45:40	
J3821-10	S	Reactive CN	P	1.047	µg/l	7/25/2018 14:02:09	
CCV3	S	Reactive CN	P	249.7299	µg/l	7/25/2018 14:02:10	
CCB3	S	Reactive CN	P	1.3108	µg/l	7/25/2018 14:02:11	