

2B96892

Test results

Aquakem 7.2AQL

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/26/2018 11:38

Test: *Reactive*
~~Total~~ CN *Mg/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.413	0.0	0.094	
ICB1	0.025	0.0	0.004	
CCV1	240.489	0.0	0.226	
CCB1	-0.106	0.0	0.004	
PB111488BL	-0.541	0.0	0.003	
J4152-04	-0.723	0.0	0.003	
J4152-04DUP	-0.321	0.0	0.004	
J4152-08	-0.327	0.0	0.004	
J4162-07	-2.616	0.0	0.002	
PB111489BL	-0.509	0.0	0.003	
J4158-02	-0.569	0.0	0.003	
J4158-02DUP	-0.529	0.0	0.003	
J4167-02	-0.666	0.0	0.003	
CCV2	258.584	0.0	0.243	
CCB2	-0.067	0.0	0.004	

N	15
Mean	39.302
SD	89.0541
CV%	226.59

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Jul 26 11:07:29 2018

Thu Jul 26 11:37:46 2018

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.7474	µg/l	7/26/2018 11:07:29	
5.0PPBCN	A	Reactive CN	P	3.8398	µg/l	7/26/2018 11:07:30	
10PPBCN	A	Reactive CN	P	9.4456	µg/l	7/26/2018 11:07:31	
50PPBCN	A	Reactive CN	P	50.1334	µg/l	7/26/2018 11:07:32	
100PPBCN	A	Reactive CN	P	101.8269	µg/l	7/26/2018 11:07:33	
250PPBCN	A	Reactive CN	P	251.7156	µg/l	7/26/2018 11:07:34	
500PPBCN	A	Reactive CN	P	498.7862	µg/l	7/26/2018 11:07:35	
ICV1	S	Reactive CN	P	97.4128	µg/l	7/26/2018 11:33:24	
ICB1	S	Reactive CN	P	0.0251	µg/l	7/26/2018 11:33:25	
CCV1	S	Reactive CN	P	240.4888	µg/l	7/26/2018 11:33:26	
CCB1	S	Reactive CN	P	-0.106	µg/l	7/26/2018 11:33:27	
PB111488BL	S	Reactive CN	P	-0.5405	µg/l	7/26/2018 11:33:28	
J4152-04	S	Reactive CN	P	-0.7231	µg/l	7/26/2018 11:33:29	
J4152-04DUP	S	Reactive CN	P	-0.3213	µg/l	7/26/2018 11:33:30	
J4152-08	S	Reactive CN	P	-0.3267	µg/l	7/26/2018 11:33:31	
J4162-07	S	Reactive CN	P	-2.6165	µg/l	7/26/2018 11:33:32	
PB111489BL	S	Reactive CN	P	-0.5086	µg/l	7/26/2018 11:33:33	
J4158-02	S	Reactive CN	P	-0.5694	µg/l	7/26/2018 11:33:34	
J4158-02DUP	S	Reactive CN	P	-0.5294	µg/l	7/26/2018 11:37:43	
J4167-02	S	Reactive CN	P	-0.6664	µg/l	7/26/2018 11:37:44	
CCV2	S	Reactive CN	P	258.584	µg/l	7/26/2018 11:37:45	
CCB2	S	Reactive CN	P	-0.067	µg/l	7/26/2018 11:37:46	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

7/26/2018 11:08


Test Total CN

Accepted 7/26/2018 11:08

Factor 1083
Bias 0.004

Coeff. of det. 0.999951

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.7474	0.0000	
2	5.0PPBCN	0.007	3.8398	5.0000	
3	10PPBCN	0.013	9.4456	10.0000	
4	50PPBCN	0.050	50.1334	50.0000	
5	100PPBCN	0.098	101.8269	100.0000	
6	250PPBCN	0.236	251.7156	250.0000	
7	500PPBCN	0.464	498.7862	500.0000	