

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/2/2018 11:17

Test: Total CN *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.049	0.0	0.096	
ICB1	-0.449	0.0	0.004	
CCV1	243.933	0.0	0.231	
CCB1	-0.344	0.0	0.004	
PB111680BL	-0.872	0.0	0.004	
PB111680BS	99.337	0.0	0.097	
LOWPB111680	9.587	0.0	0.014	
HIGHPB111680	490.335	0.0	0.460	
J4260-01	0.649	0.0	0.005	
J4260-01DUP	-0.176	0.0	0.004	
J4260-01MS	31.959	0.0	0.034	
J4260-01MSD	32.179	0.0	0.034	
J4285-01	-0.770	0.0	0.004	
J4285-05	-0.005	0.0	0.005	
CCV2	244.056	0.0	0.231	
CCB2	-0.681	0.0	0.004	
J4285-13	2.348	0.0	0.007	
J4285-17	0.326	0.0	0.005	
J4285-21	2.403	0.0	0.007	
J4285-09	0.763	0.0	0.005	
CCV3	246.538	0.0	0.233	
CCB3	-0.952	0.0	0.004	

N 22
Mean 68.101
SD 127.2218
CV% 186.81

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : *AS*

8/2/2018 10:38


Test Total CN

Accepted 8/2/2018 10:38

Factor 1077
 Bias 0.005

Coeff. of det. 0.999859

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors

1	0.0PPBCN	0.004	-0.4039	0.0000	
2	5.0PPBCN	0.008	3.3548	5.0000	
3	10PPBCN	0.013	9.2637	10.0000	
4	50PPBCN	0.050	49.1281	50.0000	
5	100PPBCN	0.099	101.9304	100.0000	
6	250PPBCN	0.240	253.9891	250.0000	
7	500PPBCN	0.467	497.7377	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Thu Aug 02 09:46:01 2018

Thu Aug 02 11:15:32 2018

AS

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.4039	µg/l	8/2/2018 10:37:56	
5.OPPBCN	A	Total CN	P	3.3548	µg/l	8/2/2018 10:37:57	
10PPBCN	A	Total CN	P	9.2637	µg/l	8/2/2018 10:37:58	
50PPBCN	A	Total CN	P	49.1281	µg/l	8/2/2018 10:37:59	
100PPBCN	A	Total CN	P	101.9304	µg/l	8/2/2018 10:38:00	
250PPBCN	A	Total CN	P	253.9891	µg/l	8/2/2018 10:38:01	
500PPBCN	A	Total CN	P	497.7377	µg/l	8/2/2018 10:38:02	
ICV1	S	Total CN	P	98.0488	µg/l	8/2/2018 11:04:33	
ICB1	S	Total CN	P	-0.449	µg/l	8/2/2018 11:04:34	
CCV1	S	Total CN	P	243.9334	µg/l	8/2/2018 11:04:35	
CCB1	S	Total CN	P	-0.3435	µg/l	8/2/2018 11:04:36	
PB111680BL	S	Total CN	P	-0.8719	µg/l	8/2/2018 11:04:37	
PB111680BS	S	Total CN	P	99.3372	µg/l	8/2/2018 11:04:38	
LOWPB111680	S	Total CN	P	9.5868	µg/l	8/2/2018 11:04:39	
HIGHPB111680	S	Total CN	P	490.3354	µg/l	8/2/2018 11:04:40	
J4260-01	S	Total CN	P	0.649	µg/l	8/2/2018 11:04:41	
J4260-01DUP	S	Total CN	P	-0.1763	µg/l	8/2/2018 11:04:42	
J4260-01MS	S	Total CN	P	31.9588	µg/l	8/2/2018 11:04:43	
J4260-01MSD	S	Total CN	P	32.1787	µg/l	8/2/2018 11:12:08	
J4285-01	S	Total CN	P	-0.7703	µg/l	8/2/2018 11:12:09	
J4285-05	S	Total CN	P	-0.0051	µg/l	8/2/2018 11:12:10	
CCV2	S	Total CN	P	244.0558	µg/l	8/2/2018 11:12:11	
CCB2	S	Total CN	P	-0.6814	µg/l	8/2/2018 11:12:12	
J4285-13	S	Total CN	P	2.3476	µg/l	8/2/2018 11:12:13	
J4285-17	S	Total CN	P	0.3263	µg/l	8/2/2018 11:12:14	
J4285-21	S	Total CN	P	2.4028	µg/l	8/2/2018 11:12:15	
J4285-09	S	Total CN	P	0.7632	µg/l	8/2/2018 11:12:16	
CCV3	S	Total CN	P	246.5376	µg/l	8/2/2018 11:12:17	
CCB3	S	Total CN	P	-0.9524	µg/l	8/2/2018 11:12:18	