

2B97193

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/9/2018 11:21

Test: ^{Reactive} Total CN ^{ugld}

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.504	0.0	0.092	
ICB1	-0.013	0.0	0.004	
CCV1	242.299	0.0	0.223	
CCB1	0.397	0.0	0.004	
PB111900BL	0.110	0.0	0.004	
J4271-02	0.108	0.0	0.004	
J4271-02DUP	0.078	0.0	0.004	
J4271-04	0.050	0.0	0.004	
J4271-06	0.169	0.0	0.004	
J4392-01	0.373	0.0	0.004	
CCV2	242.635	0.0	0.224	
CCB2	-0.360	0.0	0.004	

N	12
Mean	48.612
SD	94.7389
CV%	194.89

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/9/2018 9:47


Test Total CN

Accepted 8/9/2018 9:47

Factor 1105
Bias 0.004

Coeff. of det. 0.999961

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.9770	0.0000	
2	5.0PPBCN	0.008	4.2668	5.0000	
3	10PPBCN	0.013	9.7740	10.0000	
4	50PPBCN	0.049	49.8185	50.0000	
5	100PPBCN	0.097	102.4468	100.0000	
6	250PPBCN	0.231	250.2606	250.0000	
7	500PPBCN	0.456	499.4104	500.0000	

Aquakem v. 7.2AQ1

AS

Results from time period:

Thu Aug 09 09:41:41 2018

Thu Aug 09 11:14:48 2018

Sample Id	Sam/Ctr/c	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.977	µg/l	8/9/2018 9:41:41	
5.0PPBCN	A	Reactive CN	P	4.2668	µg/l	8/9/2018 9:41:42	
10PPBCN	A	Reactive CN	P	9.774	µg/l	8/9/2018 9:41:43	
50PPBCN	A	Reactive CN	P	49.8185	µg/l	8/9/2018 9:41:44	
100PPBCN	A	Reactive CN	P	102.4468	µg/l	8/9/2018 9:41:45	
250PPBCN	A	Reactive CN	P	250.2606	µg/l	8/9/2018 9:41:46	
500PPBCN	A	Reactive CN	P	499.4104	µg/l	8/9/2018 9:41:47	
ICV1	S	Reactive CN	P	97.5042	µg/l	8/9/2018 11:12:00	
ICB1	S	Reactive CN	P	-0.0132	µg/l	8/9/2018 11:12:01	
CCV1	S	Reactive CN	P	242.2987	µg/l	8/9/2018 11:12:02	
CCB1	S	Reactive CN	P	0.3972	µg/l	8/9/2018 11:12:03	
PB111900BL	S	Reactive CN	P	0.1097	µg/l	8/9/2018 11:12:04	
J4271-02	S	Reactive CN	P	0.108	µg/l	8/9/2018 11:12:05	
J4271-02DUP	S	Reactive CN	P	0.0779	µg/l	8/9/2018 11:12:06	
J4271-04	S	Reactive CN	P	0.0504	µg/l	8/9/2018 11:12:07	
J4271-06	S	Reactive CN	P	0.1693	µg/l	8/9/2018 11:12:08	
J4392-01	S	Reactive CN	P	0.3726	µg/l	8/9/2018 11:12:09	
CCV2	S	Reactive CN	P	242.6352	µg/l	8/9/2018 11:12:10	
CCB2	S	Reactive CN	P	-0.3602	µg/l	8/9/2018 11:14:48	