

KB97614

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/29/2018 11:24

Test: Reactive
Total CN, ug/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.801	0.0	0.098	
ICB1	0.172	0.0	0.004	
CCV1	241.347	0.0	0.241	
CCB1	0.144	0.0	0.004	
PB112469BL	-0.243	0.0	0.003	
J4663-02	-0.252	0.0	0.003	
J4663-02DUP	-0.351	0.0	0.003	
J4677-01	-0.123	0.0	0.003	
J4677-02	-0.306	0.0	0.003	
J4677-03	-0.067	0.0	0.004	
CCV2	239.487	0.0	0.239	
CCB2	-0.157	0.0	0.003	

N	12
Mean	47.954
SD	93.9948
CV%	196.01

Aquakem v. 7.2AQ1

Results from time period:

Wed Aug 29 11:21:56 2018

Wed Aug 29 11:24:44 2018

AS

Sample Id	Sam/Ctr/c	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.3942	µg/l	8/29/2018 8:11:46	
5.0PPBCN	A	Reactive CN	P	4.6169	µg/l	8/29/2018 8:11:47	
10PPBCN	A	Reactive CN	P	9.8342	µg/l	8/29/2018 8:11:48	
50PPBCN	A	Reactive CN	P	49.8586	µg/l	8/29/2018 8:11:49	
100PPBCN	A	Reactive CN	P	101.1084	µg/l	8/29/2018 8:11:50	
250PPBCN	A	Reactive CN	P	250.3531	µg/l	8/29/2018 8:11:51	
500PPBCN	A	Reactive CN	P	499.623	µg/l	8/29/2018 8:11:52	
ICV1	S	Reactive CN	P	95.8013	µg/l	8/29/2018 11:21:56	
ICB1	S	Reactive CN	P	0.1724	µg/l	8/29/2018 11:21:57	
CCV1	S	Reactive CN	P	241.3469	µg/l	8/29/2018 11:21:58	
CCB1	S	Reactive CN	P	0.1439	µg/l	8/29/2018 11:21:59	
PB112469BL	S	Reactive CN	P	-0.243	µg/l	8/29/2018 11:22:00	
J4663-02	S	Reactive CN	P	-0.2522	µg/l	8/29/2018 11:22:01	
J4663-02DUP	S	Reactive CN	P	-0.3509	µg/l	8/29/2018 11:22:02	
J4677-01	S	Reactive CN	P	-0.1229	µg/l	8/29/2018 11:22:03	
J4677-02	S	Reactive CN	P	-0.3063	µg/l	8/29/2018 11:22:04	
J4677-03	S	Reactive CN	P	-0.067	µg/l	8/29/2018 11:22:05	
CCV2	S	Reactive CN	P	239.4873	µg/l	8/29/2018 11:22:06	
CCB2	S	Reactive CN	P	-0.1567	µg/l	8/29/2018 11:24:44	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/29/2018 8:16


Test Total CN

Accepted 8/29/2018 8:16

Factor 1017
Bias 0.004

Coeff. of det. 0.999991

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.3942	0.0000	
2	5.0PPBCN	0.008	4.6169	5.0000	
3	10PPBCN	0.013	9.8342	10.0000	
4	50PPBCN	0.053	49.8586	50.0000	
5	100PPBCN	0.103	101.1084	100.0000	
6	250PPBCN	0.250	250.3531	250.0000	
7	500PPBCN	0.495	499.6230	500.0000	