

LB97632

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/30/2018 11:56

Test: *Reactive*
~~Total~~ CN *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.536	0.0	0.100	
ICB1	0.261	0.0	0.004	
CCV1	239.643	0.0	0.242	
CCB1	0.116	0.0	0.004	
PB112493BL	-0.162	0.0	0.003	
J4275-20	0.028	0.0	0.004	
J4275-20DUP	0.020	0.0	0.004	
J4275-22	0.020	0.0	0.004	
J4282-16	-0.255	0.0	0.003	
J4282-18	-0.249	0.0	0.003	
J4282-20	-0.558	0.0	0.003	
J4699-03	-0.755	0.0	0.003	
J4699-06	-0.195	0.0	0.003	
CCV2	255.582	0.0	0.258	
CCB2	-0.081	0.0	0.004	

N	15
Mean	39.330
SD	88.1836
CV%	224.21

Aquakem v. 7.2AQ1

Results from time period:

Thu Aug 30 10:59:45 2018

Thu Aug 30 11:34:07 2018

AS

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.2375	µg/l	8/30/2018 10:59:45	
5.OPPBCN	A	Reactive CN	P	4.8414	µg/l	8/30/2018 10:59:46	
10PPBCN	A	Reactive CN	P	10.0202	µg/l	8/30/2018 10:59:47	
50PPBCN	A	Reactive CN	P	50.5164	µg/l	8/30/2018 10:59:48	
100PPBCN	A	Reactive CN	P	99.9831	µg/l	8/30/2018 10:59:49	
250PPBCN	A	Reactive CN	P	249.8469	µg/l	8/30/2018 10:59:50	
500PPBCN	A	Reactive CN	P	500.0294	µg/l	8/30/2018 10:59:51	
ICV1	S	Reactive CN	P	96.5362	µg/l	8/30/2018 11:29:45	
ICB1	S	Reactive CN	P	0.2611	µg/l	8/30/2018 11:29:46	
CCV1	S	Reactive CN	P	239.6433	µg/l	8/30/2018 11:29:47	
CCB1	S	Reactive CN	P	0.1159	µg/l	8/30/2018 11:29:48	
PB112493BL	S	Reactive CN	P	-0.1615	µg/l	8/30/2018 11:29:49	
J4275-20	S	Reactive CN	P	0.0281	µg/l	8/30/2018 11:29:50	
J4275-20DUP	S	Reactive CN	P	0.0197	µg/l	8/30/2018 11:29:51	
J4275-22	S	Reactive CN	P	0.0197	µg/l	8/30/2018 11:29:52	
J4282-16	S	Reactive CN	P	-0.2553	µg/l	8/30/2018 11:29:53	
J4282-18	S	Reactive CN	P	-0.2488	µg/l	8/30/2018 11:29:54	
J4282-20	S	Reactive CN	P	-0.5584	µg/l	8/30/2018 11:29:55	
J4699-03	S	Reactive CN	P	-0.755	µg/l	8/30/2018 11:34:04	
J4699-06	S	Reactive CN	P	-0.1946	µg/l	8/30/2018 11:34:05	
CCV2	S	Reactive CN	P	255.5824	µg/l	8/30/2018 11:34:06	
CCB2	S	Reactive CN	P	-0.0814	µg/l	8/30/2018 11:34:07	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

8/30/2018 11:01


Test Total CN

Accepted 8/30/2018 11:01

Factor 1005
 Bias 0.004

Coeff. of det. 0.999998

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.2375	0.0000	
2	5.0PPBCN	0.008	4.8414	5.0000	
3	10PPBCN	0.014	10.0202	10.0000	
4	50PPBCN	0.054	50.5164	50.0000	
5	100PPBCN	0.103	99.9831	100.0000	
6	250PPBCN	0.252	249.8469	250.0000	
7	500PPBCN	0.501	500.0294	500.0000	