

LB97708

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/4/2018 13:55

Test: ^{Acetone} Total CN ug/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.692	0.0	0.097	
ICB1	0.780	0.0	0.004	
CCV1	236.526	0.0	0.235	
CCB1	0.813	0.0	0.004	
PB112574BL	0.574	0.0	0.004	
J4747-04	0.409	0.0	0.003	
J4747-04DUP	0.361	0.0	0.003	
J4756-03	0.495	0.0	0.003	
J4756-04	0.674	0.0	0.004	
J4757-02	0.407	0.0	0.003	
J4761-03	0.504	0.0	0.003	
J4761-06	0.532	0.0	0.003	
J4764-02	0.738	0.0	0.004	
CCV2	253.234	0.0	0.251	
CCB2	0.351	0.0	0.003	

N	15
Mean	39.473
SD	86.9570
CV%	220.30

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Sep 04 13:48:49 2018

Tue Sep 04 13:53:11 2018

Sample Id	Sam/Ctr/c	Test short narr	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.3636	µg/l	9/4/2018 10:16:42	
5.OPPBCN	A	Reactive CN	P	4.9527	µg/l	9/4/2018 10:16:43	
10PPBCN	A	Reactive CN	P	10.3965	µg/l	9/4/2018 10:16:44	
50PPBCN	A	Reactive CN	P	50.4276	µg/l	9/4/2018 10:16:45	
100PPBCN	A	Reactive CN	P	100.4026	µg/l	9/4/2018 10:16:46	
250PPBCN	A	Reactive CN	P	247.1754	µg/l	9/4/2018 10:16:47	
500PPBCN	A	Reactive CN	P	501.2816	µg/l	9/4/2018 10:16:48	
ICV1	S	Reactive CN	P	95.6919	µg/l	9/4/2018 13:48:49	
ICB1	S	Reactive CN	P	0.7795	µg/l	9/4/2018 13:48:50	
CCV1	S	Reactive CN	P	236.5257	µg/l	9/4/2018 13:48:51	
CCB1	S	Reactive CN	P	0.8132	µg/l	9/4/2018 13:48:52	
PB112574BL	S	Reactive CN	P	0.5742	µg/l	9/4/2018 13:48:53	
J4747-04	S	Reactive CN	P	0.4085	µg/l	9/4/2018 13:48:54	
J4747-04DUP	S	Reactive CN	P	0.3609	µg/l	9/4/2018 13:48:55	
J4756-03	S	Reactive CN	P	0.495	µg/l	9/4/2018 13:48:56	
J4756-04	S	Reactive CN	P	0.674	µg/l	9/4/2018 13:48:57	
J4757-02	S	Reactive CN	P	0.4073	µg/l	9/4/2018 13:48:58	
J4761-03	S	Reactive CN	P	0.5036	µg/l	9/4/2018 13:48:59	
J4761-06	S	Reactive CN	P	0.5324	µg/l	9/4/2018 13:53:08	
J4764-02	S	Reactive CN	P	0.7382	µg/l	9/4/2018 13:53:09	
CCV2	S	Reactive CN	P	253.2336	µg/l	9/4/2018 13:53:10	
CCB2	S	Reactive CN	P	0.351	µg/l	9/4/2018 13:53:11	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/4/2018 10:25


Test Total CN

Accepted 9/4/2018 10:25

Factor 1021
Bias 0.003

Coeff. of det. 0.999950

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.3636	0.0000	
2	5.0PPBCN	0.008	4.9527	5.0000	
3	10PPBCN	0.013	10.3965	10.0000	
4	50PPBCN	0.052	50.4276	50.0000	
5	100PPBCN	0.101	100.4026	100.0000	
6	250PPBCN	0.245	247.1754	250.0000	
7	500PPBCN	0.494	501.2816	500.0000	