

1397728

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/5/2018 12:20

Test: *Reactive* Total CN *Mg/L*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.789	0.0	0.102	
ICB1	-0.476	0.0	0.004	
CCV1	247.176	0.0	0.250	
CCB1	-0.429	0.0	0.004	
PB112626BL	-0.740	0.0	0.003	
J4781-02	-0.683	0.0	0.004	
J4781-02DUP	-0.859	0.0	0.003	
CCV2	247.331	0.0	0.250	
CCB2	-0.379	0.0	0.004	

N	9
Mean	65.526
SD	108.0438
CV%	164.89

21397728

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Sep 05 11:52:17 2018

Wed Sep 05 11:52:25 2018

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.8249	µg/l	9/5/2018 8:52:50	
5.OPPBCN	A	Reactive CN	P	4.0507	µg/l	9/5/2018 8:52:51	
10PPBCN	A	Reactive CN	P	9.3111	µg/l	9/5/2018 8:52:52	
50PPBCN	A	Reactive CN	P	49.207	µg/l	9/5/2018 8:52:53	
100PPBCN	A	Reactive CN	P	102.4228	µg/l	9/5/2018 8:52:54	
250PPBCN	A	Reactive CN	P	252.4305	µg/l	9/5/2018 8:52:55	
500PPBCN	A	Reactive CN	P	498.4027	µg/l	9/5/2018 8:52:56	
ICV1	S	Reactive CN	P	98.7886	µg/l	9/5/2018 11:52:17	
ICB1	S	Reactive CN	P	-0.476	µg/l	9/5/2018 11:52:18	
CCV1	S	Reactive CN	P	247.1761	µg/l	9/5/2018 11:52:19	
CCB1	S	Reactive CN	P	-0.4286	µg/l	9/5/2018 11:52:20	
PB112626BL	S	Reactive CN	P	-0.7396	µg/l	9/5/2018 11:52:21	
J4781-02	S	Reactive CN	P	-0.6828	µg/l	9/5/2018 11:52:22	
J4781-02DUP	S	Reactive CN	P	-0.8592	µg/l	9/5/2018 11:52:23	
CCV2	S	Reactive CN	P	247.3311	µg/l	9/5/2018 11:52:24	
CCB2	S	Reactive CN	P	-0.3786	µg/l	9/5/2018 11:52:25	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/5/2018 8:53


Test Total CN

Accepted 9/5/2018 8:53

Factor 1008
Bias 0.004

Coeff. of det. 0.999917

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.8249	0.0000	
2	5.0PPBCN	0.008	4.0507	5.0000	
3	10PPBCN	0.013	9.3111	10.0000	
4	50PPBCN	0.053	49.2070	50.0000	
5	100PPBCN	0.106	102.4228	100.0000	
6	250PPBCN	0.255	252.4305	250.0000	
7	500PPBCN	0.499	498.4027	500.0000	