

AB91759

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/6/2018 14:15

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.885	0.0	0.092	
ICB1	0.641	0.0	0.004	
CCV1	231.353	0.0	0.218	
CCB1	0.629	0.0	0.004	
PB112662BL	0.355	0.0	0.004	
PB112662BS	98.112	0.0	0.094	
MIDPB112662	234.671	0.0	0.221	
J4799-01	1.346	0.0	0.005	
J4799-01DUP	0.809	0.0	0.004	
J4799-01MS	31.796	0.0	0.033	
J4799-01MSD	31.753	0.0	0.033	
J4806-01	0.336	0.0	0.004	
J4806-02	1.379	0.0	0.005	
J4806-03	0.431	0.0	0.004	
CCV2	245.871	0.0	0.232	
CCB2	0.298	0.0	0.004	

N 16
Mean 60.979
SD 93.2369
CV% 152.90

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/6/2018 12:52


Test Total CN

Accepted 9/6/2018 12:51

Factor 1077
Bias 0.003

Coeff. of det. 0.999740

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.1764	0.0000	
2	5.0PPBCN	0.008	5.2231	5.0000	
3	10PPBCN	0.013	10.4181	10.0000	
4	50PPBCN	0.051	50.7833	50.0000	
5	100PPBCN	0.099	102.4938	100.0000	
6	250PPBCN	0.230	243.6918	250.0000	
7	500PPBCN	0.470	502.5664	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Thu Sep 06 14:06:41 2018

Thu Sep 06 14:11:32 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.1764	µg/l	9/6/2018 12:49:05	
5.OPPBCN	A	Total CN	P	5.2231	µg/l	9/6/2018 12:49:06	
10PPBCN	A	Total CN	P	10.4181	µg/l	9/6/2018 12:49:07	
50PPBCN	A	Total CN	P	50.7833	µg/l	9/6/2018 12:49:08	
100PPBCN	A	Total CN	P	102.4938	µg/l	9/6/2018 12:49:09	
250PPBCN	A	Total CN	P	243.6918	µg/l	9/6/2018 12:49:10	
500PPBCN	A	Total CN	P	502.5664	µg/l	9/6/2018 12:49:11	
ICV1	S	Total CN	P	95.8852	µg/l	9/6/2018 14:06:41	
ICB1	S	Total CN	P	0.6412	µg/l	9/6/2018 14:06:42	
CCV1	S	Total CN	P	231.3526	µg/l	9/6/2018 14:06:43	
CCB1	S	Total CN	P	0.6292	µg/l	9/6/2018 14:06:44	
PB112662BL	S	Total CN	P	0.355	µg/l	9/6/2018 14:06:45	
PB112662BS	S	Total CN	P	98.1119	µg/l	9/6/2018 14:06:46	
MIDPB112662	S	Total CN	P	234.6707	µg/l	9/6/2018 14:06:47	
J4799-01	S	Total CN	P	1.346	µg/l	9/6/2018 14:06:48	
J4799-01DUP	S	Total CN	P	0.8086	µg/l	9/6/2018 14:06:49	
J4799-01MS	S	Total CN	P	31.796	µg/l	9/6/2018 14:06:50	
J4799-01MSD	S	Total CN	P	31.7533	µg/l	9/6/2018 14:06:51	
J4806-01	S	Total CN	P	0.3355	µg/l	9/6/2018 14:11:28	
J4806-02	S	Total CN	P	1.3786	µg/l	9/6/2018 14:11:29	
J4806-03	S	Total CN	P	0.4306	µg/l	9/6/2018 14:11:30	
CCV2	S	Total CN	P	245.8709	µg/l	9/6/2018 14:11:31	
CCB2	S	Total CN	P	0.2977	µg/l	9/6/2018 14:11:32	