

2B97884

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/13/2018 12:33

Test: Total CN My/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.587	0.0	0.092	
ICB1	0.444	0.0	0.004	
CCV1	242.287	0.0	0.229	
CCB1	0.609	0.0	0.004	
PB112810BL	0.425	0.0	0.004	
PB112810BS	96.443	0.0	0.093	
LOWPB112810	10.466	0.0	0.013	
HIGHPB112810	492.537	0.0	0.462	
J4771-01	6.076	0.0	0.009	
J4771-01DUP	5.322	0.0	0.008	
J4771-01MS	36.727	0.0	0.038	
J4771-01MSD	36.702	0.0	0.038	
J4873-01	1.715	0.0	0.005	
J4873-03	1.152	0.0	0.005	
CCV2	245.081	0.0	0.232	
CCB2	0.386	0.0	0.004	
J4873-05	0.916	0.0	0.004	
J4883-01	2.342	0.0	0.006	
PB112811BL	0.111	0.0	0.004	
PB112811BS	97.010	0.0	0.094	
J4882-01	9.571	0.0	0.012	
J4882-01DUP	9.454	0.0	0.012	
J4882-01MS	41.764	0.0	0.042	
J4882-01MSD	42.342	0.0	0.043	
J4883-03	1.057	0.0	0.004	
CCV3	246.620	0.0	0.233	
CCB3	0.687	0.0	0.004	

N	27
Mean	63.846
SD	115.4545
CV%	180.83

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/13/2018 10:02


Test Total CN

Accepted 9/13/2018 10:02

Factor 1074
Bias 0.003

Coeff. of det. 0.999975

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.0584	0.0000	
2	5.0PPBCN	0.008	4.7940	5.0000	
3	10PPBCN	0.013	9.9743	10.0000	
4	50PPBCN	0.050	49.8247	50.0000	
5	100PPBCN	0.096	99.3268	100.0000	
6	250PPBCN	0.238	251.9678	250.0000	
7	500PPBCN	0.468	499.1708	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Thu Sep 13 12:19:27 2018

Thu Sep 13 12:33:20 2018



Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.0584	µg/l	9/13/2018 9:58:58	
5.0PPBCN	A	Total CN	P	4.794	µg/l	9/13/2018 9:58:59	
10PPBCN	A	Total CN	P	9.9743	µg/l	9/13/2018 9:59:00	
50PPBCN	A	Total CN	P	49.8247	µg/l	9/13/2018 9:59:01	
100PPBCN	A	Total CN	P	99.3268	µg/l	9/13/2018 9:59:02	
250PPBCN	A	Total CN	P	251.9678	µg/l	9/13/2018 9:59:03	
500PPBCN	A	Total CN	P	499.1708	µg/l	9/13/2018 9:59:04	
ICV1	S	Total CN	P	95.5875	µg/l	9/13/2018 12:19:27	
ICB1	S	Total CN	P	0.444	µg/l	9/13/2018 12:19:28	
CCV1	S	Total CN	P	242.2875	µg/l	9/13/2018 12:19:29	
CCB1	S	Total CN	P	0.6093	µg/l	9/13/2018 12:19:30	
PB112810BL	S	Total CN	P	0.4247	µg/l	9/13/2018 12:19:31	
PB112810BS	S	Total CN	P	96.4429	µg/l	9/13/2018 12:19:32	
LOWPB112810	S	Total CN	P	10.466	µg/l	9/13/2018 12:19:33	
HIGHPB112810	S	Total CN	P	492.5368	µg/l	9/13/2018 12:19:34	
J4771-01	S	Total CN	P	6.0758	µg/l	9/13/2018 12:19:35	
J4771-01DUP	S	Total CN	P	5.3223	µg/l	9/13/2018 12:19:36	
J4771-01MS	S	Total CN	P	36.7265	µg/l	9/13/2018 12:19:37	
J4771-01MSD	S	Total CN	P	36.702	µg/l	9/13/2018 12:27:02	
J4873-01	S	Total CN	P	1.7154	µg/l	9/13/2018 12:27:03	
J4873-03	S	Total CN	P	1.1517	µg/l	9/13/2018 12:27:04	
CCV2	S	Total CN	P	245.0806	µg/l	9/13/2018 12:27:05	
CCB2	S	Total CN	P	0.3856	µg/l	9/13/2018 12:27:06	
J4873-05	S	Total CN	P	0.9161	µg/l	9/13/2018 12:27:07	
J4883-01	S	Total CN	P	2.3415	µg/l	9/13/2018 12:27:08	
PB112811BL	S	Total CN	P	0.1106	µg/l	9/13/2018 12:27:09	
PB112811BS	S	Total CN	P	97.0097	µg/l	9/13/2018 12:27:10	
J4882-01	S	Total CN	P	9.5711	µg/l	9/13/2018 12:27:11	
J4882-01DUP	S	Total CN	P	9.4538	µg/l	9/13/2018 12:27:12	
J4882-01MS	S	Total CN	P	41.7644	µg/l	9/13/2018 12:33:13	
J4882-01MSD	S	Total CN	P	42.3421	µg/l	9/13/2018 12:33:14	
J4883-03	S	Total CN	P	1.0568	µg/l	9/13/2018 12:33:15	
CCV3	S	Total CN	P	246.6196	µg/l	9/13/2018 12:33:16	
CCB3	S	Total CN	P	0.6873	µg/l	9/13/2018 12:33:17	