

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/13/2018 11:43

Test: *Reactive* Total CN *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.798	0.0	0.093	
ICB1	0.520	0.0	0.004	
CCV1	241.824	0.0	0.229	
CCB1	0.543	0.0	0.004	
PB112806BL	0.290	0.0	0.004	
J4771-02	0.328	0.0	0.004	
J4771-02DUP	0.376	0.0	0.004	
J4873-02	0.226	0.0	0.004	
J4873-04	0.164	0.0	0.004	
J4873-06	0.453	0.0	0.004	
J4883-02	0.090	0.0	0.004	
PB112807BL	0.080	0.0	0.004	
J4882-02	0.270	0.0	0.004	
J4882-02DUP	0.525	0.0	0.004	
CCV2	250.041	0.0	0.236	
CCB2	0.480	0.0	0.004	
J4883-04	0.212	0.0	0.004	
CCV3	250.512	0.0	0.237	
CCB3	0.282	0.0	0.004	

N	19
Mean	44.369
SD	92.9549
CV%	209.50

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/13/2018 10:02


Test Total CN

Accepted 9/13/2018 10:02

Factor 1074
Bias 0.003

Coeff. of det. 0.999975

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.0584	0.0000	
2	5.0PPBCN	0.008	4.7940	5.0000	
3	10PPBCN	0.013	9.9743	10.0000	
4	50PPBCN	0.050	49.8247	50.0000	
5	100PPBCN	0.096	99.3268	100.0000	
6	250PPBCN	0.238	251.9678	250.0000	
7	500PPBCN	0.468	499.1708	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Thu Sep 13 09:58:58 2018

Thu Sep 13 11:42:07 2018

AS

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.0584	µg/l	9/13/2018 9:58:58	
5.OPPBCN	A	Reactive CN	P	4.794	µg/l	9/13/2018 9:58:59	
10PPBCN	A	Reactive CN	P	9.9743	µg/l	9/13/2018 9:59:00	
50PPBCN	A	Reactive CN	P	49.8247	µg/l	9/13/2018 9:59:01	
100PPBCN	A	Reactive CN	P	99.3268	µg/l	9/13/2018 9:59:02	
250PPBCN	A	Reactive CN	P	251.9678	µg/l	9/13/2018 9:59:03	
500PPBCN	A	Reactive CN	P	499.1708	µg/l	9/13/2018 9:59:04	
ICV1	S	Reactive CN	P	95.7979	µg/l	9/13/2018 11:35:49	
ICB1	S	Reactive CN	P	0.5197	µg/l	9/13/2018 11:35:50	
CCV1	S	Reactive CN	P	241.8241	µg/l	9/13/2018 11:35:51	
CCB1	S	Reactive CN	P	0.543	µg/l	9/13/2018 11:35:52	
PB112806BL	S	Reactive CN	P	0.2901	µg/l	9/13/2018 11:35:53	
J4771-02	S	Reactive CN	P	0.3282	µg/l	9/13/2018 11:35:54	
J4771-02DUP	S	Reactive CN	P	0.376	µg/l	9/13/2018 11:35:55	
J4873-02	S	Reactive CN	P	0.2259	µg/l	9/13/2018 11:35:56	
J4873-04	S	Reactive CN	P	0.1642	µg/l	9/13/2018 11:35:57	
J4873-06	S	Reactive CN	P	0.4532	µg/l	9/13/2018 11:35:58	
J4883-02	S	Reactive CN	P	0.0898	µg/l	9/13/2018 11:35:59	
PB112807BL	S	Reactive CN	P	0.0797	µg/l	9/13/2018 11:42:00	
J4882-02	S	Reactive CN	P	0.2699	µg/l	9/13/2018 11:42:01	
J4882-02DUP	S	Reactive CN	P	0.5252	µg/l	9/13/2018 11:42:02	
CCV2	S	Reactive CN	P	250.041	µg/l	9/13/2018 11:42:03	
CCB2	S	Reactive CN	P	0.4804	µg/l	9/13/2018 11:42:04	
J4883-04	S	Reactive CN	P	0.2124	µg/l	9/13/2018 11:42:05	
CCV3	S	Reactive CN	P	250.5123	µg/l	9/13/2018 11:42:06	
CCB3	S	Reactive CN	P	0.282	µg/l	9/13/2018 11:42:07	