

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/18/2018 15:03

Test: *Reactive*
~~Total~~ CN *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.148	0.0	0.093	
ICB1	0.995	0.0	0.004	
CCV1	240.916	0.0	0.223	
CCB1	0.447	0.0	0.004	
PB112965BL	0.378	0.0	0.004	
J4772-02	0.439	0.0	0.004	
J4772-02DUP	0.488	0.0	0.004	
PB112966BL	0.234	0.0	0.003	
J4929-14	1.175	0.0	0.004	
J4929-14DUP	0.536	0.0	0.004	
J4956-03	0.202	0.0	0.003	
J4957-02	0.329	0.0	0.004	
CCV2	261.638	0.0	0.242	
CCB2	0.583	0.0	0.004	

N	14
Mean	43.322
SD	91.9267
CV%	212.19

Aquakem v. 7.2AQ1

Results from time period:

Tue Sep 18 14:36:21 2018

Tue Sep 18 14:40:14 2018

AS

Sample Id	Sam/Ctr/c	Test short	nan	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive	CN	P	0.6725	µg/l	9/18/2018 11:23:36	
5.0PPBCN	A	Reactive	CN	P	5.3515	µg/l	9/18/2018 11:23:37	
10PPBCN	A	Reactive	CN	P	9.6629	µg/l	9/18/2018 11:23:38	
50PPBCN	A	Reactive	CN	P	49.5878	µg/l	9/18/2018 11:23:39	
100PPBCN	A	Reactive	CN	P	101.2554	µg/l	9/18/2018 11:23:40	
250PPBCN	A	Reactive	CN	P	247.3529	µg/l	9/18/2018 11:23:41	
500PPBCN	A	Reactive	CN	P	501.1169	µg/l	9/18/2018 11:23:42	
ICV1	S	Reactive	CN	P	98.1477	µg/l	9/18/2018 14:36:21	
ICB1	S	Reactive	CN	P	0.9953	µg/l	9/18/2018 14:36:22	
CCV1	S	Reactive	CN	P	240.9157	µg/l	9/18/2018 14:36:23	
CCB1	S	Reactive	CN	P	0.4473	µg/l	9/18/2018 14:36:24	
PB112965BL	S	Reactive	CN	P	0.378	µg/l	9/18/2018 14:36:25	
J4772-02	S	Reactive	CN	P	0.4388	µg/l	9/18/2018 14:36:26	
J4772-02DUP	S	Reactive	CN	P	0.4881	µg/l	9/18/2018 14:36:27	
PB112966BL	S	Reactive	CN	P	0.2339	µg/l	9/18/2018 14:36:28	
J4929-14	S	Reactive	CN	P	1.1751	µg/l	9/18/2018 14:36:29	
J4929-14DUP	S	Reactive	CN	P	0.5362	µg/l	9/18/2018 14:36:30	
J4956-03	S	Reactive	CN	P	0.2016	µg/l	9/18/2018 14:36:31	
J4957-02	S	Reactive	CN	P	0.3292	µg/l	9/18/2018 14:40:12	
CCV2	S	Reactive	CN	P	261.6381	µg/l	9/18/2018 14:40:13	
CCB2	S	Reactive	CN	P	0.5834	µg/l	9/18/2018 14:40:14	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/18/2018 11:23


Test Total CN

Accepted 9/18/2018 11:23

Factor 1097
Bias 0.003

Coeff. of det. 0.999948

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.6725	0.0000	
2	5.0PPBCN	0.008	5.3515	5.0000	
3	10PPBCN	0.012	9.6629	10.0000	
4	50PPBCN	0.048	49.5878	50.0000	
5	100PPBCN	0.096	101.2554	100.0000	
6	250PPBCN	0.229	247.3529	250.0000	
7	500PPBCN	0.460	501.1169	500.0000	