

21398161

=====

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/26/2018 13:04

Test: *Reactive*
~~Total~~ CN *ug/l*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.176	0.0	0.097	
ICB1	0.283	0.0	0.004	
CCV1	244.191	0.0	0.241	
CCB1	-0.080	0.0	0.004	
PB113234BL	-0.252	0.0	0.004	
J4773-12	-0.255	0.0	0.004	
J4773-12DUP	-0.289	0.0	0.004	
J4773-14	-0.253	0.0	0.004	
J4773-16	-0.254	0.0	0.004	
J4773-18	-0.105	0.0	0.004	
J4773-20	-0.063	0.0	0.004	
J5074-10	-0.026	0.0	0.004	
J5074-20	-0.378	0.0	0.004	
J5081-04	-0.100	0.0	0.004	
CCV2	249.671	0.0	0.246	
CCB2	-0.530	0.0	0.003	
J5081-08	-0.373	0.0	0.004	
J5081-12	-0.218	0.0	0.004	
J5081-16	-0.934	0.0	0.003	
CCV3	247.578	0.0	0.244	
CCB3	-0.133	0.0	0.004	
N	21			
Mean	39.698			
SD	89.2797			
CV%	224.90			

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/26/2018 9:31


Test Total CN

Accepted 9/26/2018 9:31

Factor 1030
Bias 0.004

Coeff. of det. 0.999938

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.7795	0.0000	
2	5.0PPBCN	0.008	4.3315	5.0000	
3	10PPBCN	0.013	9.5010	10.0000	
4	50PPBCN	0.052	49.6145	50.0000	
5	100PPBCN	0.102	101.1234	100.0000	
6	250PPBCN	0.249	252.7571	250.0000	
7	500PPBCN	0.488	498.4520	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Wed Sep 26 09:18:08 2018

Wed Sep 26 12:54:34 2018

✓ AS

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.7795	µg/l	9/26/2018 9:18:08	
5.OPPBCN	A	Reactive CN	P	4.3315	µg/l	9/26/2018 9:18:09	
10PPBCN	A	Reactive CN	P	9.501	µg/l	9/26/2018 9:18:10	
50PPBCN	A	Reactive CN	P	49.6145	µg/l	9/26/2018 9:18:11	
100PPBCN	A	Reactive CN	P	101.1234	µg/l	9/26/2018 9:18:12	
250PPBCN	A	Reactive CN	P	252.7571	µg/l	9/26/2018 9:18:13	
500PPBCN	A	Reactive CN	P	498.452	µg/l	9/26/2018 9:18:14	
ICV1	S	Reactive CN	P	96.1759	µg/l	9/26/2018 12:05:13	
ICB1	S	Reactive CN	P	0.2831	µg/l	9/26/2018 12:05:14	
CCV1	S	Reactive CN	P	244.1914	µg/l	9/26/2018 12:48:16	
CCB1	S	Reactive CN	P	-0.0804	µg/l	9/26/2018 12:48:17	
PB113234BL	S	Reactive CN	P	-0.2522	µg/l	9/26/2018 12:48:18	
J4773-12	S	Reactive CN	P	-0.2548	µg/l	9/26/2018 12:48:19	
J4773-12DUP	S	Reactive CN	P	-0.289	µg/l	9/26/2018 12:48:20	
J4773-14	S	Reactive CN	P	-0.2527	µg/l	9/26/2018 12:48:21	
J4773-16	S	Reactive CN	P	-0.2539	µg/l	9/26/2018 12:48:22	
J4773-18	S	Reactive CN	P	-0.1052	µg/l	9/26/2018 12:48:23	
J4773-20	S	Reactive CN	P	-0.0627	µg/l	9/26/2018 12:48:24	
J5074-10	S	Reactive CN	P	-0.0261	µg/l	9/26/2018 12:48:25	
J5074-20	S	Reactive CN	P	-0.3781	µg/l	9/26/2018 12:48:26	
J5081-04	S	Reactive CN	P	-0.1003	µg/l	9/26/2018 12:54:27	
CCV2	S	Reactive CN	P	249.6711	µg/l	9/26/2018 12:54:28	
CCB2	S	Reactive CN	P	-0.5298	µg/l	9/26/2018 12:54:29	
J5081-08	S	Reactive CN	P	-0.373	µg/l	9/26/2018 12:54:30	
J5081-12	S	Reactive CN	P	-0.2185	µg/l	9/26/2018 12:54:31	
J5081-16	S	Reactive CN	P	-0.9343	µg/l	9/26/2018 12:54:32	
CCV3	S	Reactive CN	P	247.5781	µg/l	9/26/2018 12:54:33	
CCB3	S	Reactive CN	P	-0.1334	µg/l	9/26/2018 12:54:34	