

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/1/2018 15:49

Test: Phenolics-mg/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.084	0.0	0.116	
ICB1	0.006	0.0	0.005	
CCV1	1.064	0.0	0.114	
CCB1	0.005	0.0	0.005	
PB113307BL	0.004	0.0	0.005	
PB113307BS	1.086	0.0	0.116	
J5090-01	0.005	0.0	0.005	
J5090-01DUP	0.004	0.0	0.004	
J5090-01MS	0.881	0.0	0.095	
J5090-01MSD	0.878	0.0	0.095	
J5090-02	0.003	0.0	0.004	
J5090-03	0.005	0.0	0.005	
CCV2	1.057	0.0	0.113	
CCB2	0.004	0.0	0.004	

N 14
Mean 0.435
SD 0.5191
CV% 119.42

AS
10/1/18 CCV1

21398272

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/1/2018 15:20

Test Phenolics-


Accepted 10/1/2018 15:20

Factor 9.669

Bias 0.004

Coeff. of det. 0.999987

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0523	0.0500	
2	Phenol-2	0.015	0.1011	0.1000	
3	Phenol-2	0.030	0.2501	0.2500	
4	Phenol-2	0.055	0.4947	0.5000	
5	Phenol-2	0.108	1.0013	1.0000	
6	Phenol-2	0.211	2.0006	2.0000	

Aquakem v. 7.2AQ1

Results from time period:

Mon Oct 01 15:14:51 2018

Mon Oct 01 15:48:52 2018

AS

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0523	mg/l	10/1/2018 15:14:51	
0.1PPM	A	Phenolics-	P	0.1011	mg/l	10/1/2018 15:14:52	
0.25PPM	A	Phenolics-	P	0.2501	mg/l	10/1/2018 15:14:53	
0.50PPM	A	Phenolics-	P	0.4947	mg/l	10/1/2018 15:14:54	
1.0PPM	A	Phenolics-	P	1.0013	mg/l	10/1/2018 15:14:55	
2.0PPM	A	Phenolics-	P	2.0006	mg/l	10/1/2018 15:14:56	
ICV1	S	Phenolics-	P	1.0835	mg/l	10/1/2018 15:43:55	
ICB1	S	Phenolics-	P	0.0058	mg/l	10/1/2018 15:43:56	
CCV'1	S	Phenolics-	P	1.0639	mg/l	10/1/2018 15:43:57	
CCB1	S	Phenolics-	P	0.0047	mg/l	10/1/2018 15:43:58	
PB113307BL	S	Phenolics-	P	0.0045	mg/l	10/1/2018 15:43:59	
PB113307BS	S	Phenolics-	P	1.0858	mg/l	10/1/2018 15:44:00	
J5090-01	S	Phenolics-	P	0.0053	mg/l	10/1/2018 15:44:01	
J5090-01DUP	S	Phenolics-	P	0.0042	mg/l	10/1/2018 15:44:02	
J5090-01MS	S	Phenolics-	P	0.8806	mg/l	10/1/2018 15:44:03	
J5090-01MSD	S	Phenolics-	P	0.8782	mg/l	10/1/2018 15:44:04	
J5090-02	S	Phenolics-	P	0.0034	mg/l	10/1/2018 15:44:05	
J5090-03	S	Phenolics-	P	0.0054	mg/l	10/1/2018 15:44:06	
CCV2	S	Phenolics-	P	1.057	mg/l	10/1/2018 15:48:49	
CCB2	S	Phenolics-	P	0.0036	mg/l	10/1/2018 15:48:50	