

21398538

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : ds

10/11/2018 9:11

Test: Phenolics-mg/l

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.033	0.0	0.110	
ICB1	0.003	0.0	0.004	
CCV1	0.963	0.0	0.103	
CCB1	0.006	0.0	0.004	
PB113725BL	0.004	0.0	0.004	
PB113725BS	1.039	0.0	0.111	
J5330-01	0.016	0.0	0.005	
J5360-01	0.010	0.0	0.005	
J5360-01DUP	0.010	0.0	0.005	
J5360-01MS	0.832	0.0	0.090	
J5360-01MSD	0.829	0.0	0.089	
CCV2	0.968	0.0	0.104	
CCB2	0.002	0.0	0.004	

N 13
Mean 0.440
SD 0.4896
CV% 111.35

Aquakem v. 7.2AQ1

Results from time period:

Thu Oct 11 07:51:30 2018

Thu Oct 11 08:20:29 2018

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0563	mg/l	10/11/2018 7:51:30	
0.1PPM	A	Phenolics-	P	0.1022	mg/l	10/11/2018 7:51:31	
0.25PPM	A	Phenolics-	P	0.2482	mg/l	10/11/2018 7:51:32	
0.50PPM	A	Phenolics-	P	0.5015	mg/l	10/11/2018 7:51:33	
1.0PPM	A	Phenolics-	P	0.9843	mg/l	10/11/2018 7:51:34	
2.0PPM	A	Phenolics-	P	2.0074	mg/l	10/11/2018 7:51:35	
ICV1	S	Phenolics-	P	1.0332	mg/l	10/11/2018 8:16:01	
ICB1	S	Phenolics-	P	0.0035	mg/l	10/11/2018 8:16:02	
CCV1	S	Phenolics-	P	0.9627	mg/l	10/11/2018 8:16:03	
CCB1	S	Phenolics-	P	0.0063	mg/l	10/11/2018 8:16:04	
PB113725BL	S	Phenolics-	P	0.004	mg/l	10/11/2018 8:16:05	
PB113725BS	S	Phenolics-	P	1.0393	mg/l	10/11/2018 8:16:06	
J5330-01	S	Phenolics-	P	0.0162	mg/l	10/11/2018 8:16:07	
J5360-01	S	Phenolics-	P	0.0105	mg/l	10/11/2018 8:16:08	
J5360-01DUP	S	Phenolics-	P	0.0101	mg/l	10/11/2018 8:16:09	
J5360-01MS	S	Phenolics-	P	0.832	mg/l	10/11/2018 8:16:10	
J5360-01MSD	S	Phenolics-	P	0.8286	mg/l	10/11/2018 8:16:11	
CCV2	S	Phenolics-	P	0.9677	mg/l	10/11/2018 8:16:12	
CCB2	S	Phenolics-	P	0.0023	mg/l	10/11/2018 8:20:27	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/11/2018 7:53


Test Phenolics-

Accepted 10/11/2018 7:53

Factor 9.677
Bias 0.004

Coeff. of det. 0.999875

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0563	0.0500	
2	Phenol-2	0.014	0.1022	0.1000	
3	Phenol-2	0.029	0.2482	0.2500	
4	Phenol-2	0.055	0.5015	0.5000	
5	Phenol-2	0.105	0.9843	1.0000	
6	Phenol-2	0.211	2.0074	2.0000	