

Clear table

IC-2 9056 A water

LOD/LOQ/MDL

AK 11-02-18


ident	concentrati on F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-10-30_	10/30/18 11:26		AK/AP
STD2	0.1510	0.2420	0.2640	0.7760	0.1980	0.0380	1.3610	_2018-10-30_	10/30/18 11:57		AK/AP
STD3	0.4400	0.6370	0.6380	2.1190	0.5300	1.0940	3.2410	_2018-10-30_	10/30/18 12:28		AK/AP
STD4	0.7990	1.1800	1.1760	3.9650	0.9870	2.1730	5.8600	_2018-10-30_	10/30/18 12:59		AK/AP
STD5	1.8800	2.8760	2.8510	9.5790	2.3920	5.0390	14.1420	_2018-10-30_	10/30/18 13:29		AK/AP
STD6	3.9590	5.8770	5.8660	19.6300	4.9090	9.8150	29.3570	_2018-10-30_	10/30/18 14:00		AK/AP
STD7	5.0720	7.6380	7.6550	25.4310	6.3580	12.5910	38.2890	_2018-10-30_	10/30/18 14:31		AK/AP
ICV	1.8780	2.8700	2.8610	9.5870	2.3930	5.1260	14.1540	_2018-10-30_	10/30/18 15:14		AK/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-10-30_	10/30/18 15:45		AK/AP
CCV	2.0360	2.9070	2.8930	9.7340	2.4340	5.4280	14.3430	_2018-10-30_	10/30/18 22:28		AK/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-10-30_	10/30/18 22:59		AK/AP
LB98969BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-10-30_	10/30/18 23:30		AK/AP
LB98969BSW	2.1420	2.9830	2.9680	9.9740	2.4960	5.5000	14.7100	_2018-10-31_	10/31/18 0:00		AK/AP
J5164-07	0.1260	0.1700	0.2120	0.5750	0.1500	-0.0210	1.0760	_2018-10-31_	10/31/18 0:31		AK/AP
J5164-09	0.1320	0.1750	0.2110	0.5850	0.1490	-0.0200	1.0710	_2018-10-31_	10/31/18 1:02		AK/AP
J5164-07	0.1850	0.2460	0.2730	0.8080	0.2040	0.2190	1.3780	_2018-10-31_	10/31/18 1:33		AK/AP
J5164-09	0.1750	0.2330	0.2750	0.7970	0.2010	0.2190	1.3750	_2018-10-31_	10/31/18 2:04		AK/AP
J5164-08	0.2830	0.3890	0.4100	1.3040	0.3270	0.6960	2.0580	_2018-10-31_	10/31/18 2:35		AK/AP
CCV	2.0470	2.9340	2.9100	9.7660	2.4400	5.4650	14.3840	_2018-10-31_	10/31/18 3:37		AK/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2018-10-31_	10/31/18 4:08		AK/AP

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-103018.mtw
 Equation: $Q = 0.0173947 \cdot A + 1.69339$
 RSD: 3.891 %
 Correlation coefficient: 0.999382

IC-2 cal

10-30-18 AL




K3 = 0 K2 = 0 K1 = 0.0173947 K0 = 1.69339
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	5.89	76.11	0.1	20	5.166		
3	28.61	408.3	0.4	20	5.166		
4	56.41	820.9	0.8	20	5.166		
5	148.4	2064	2	20	5.166		
6	318.2	4454	4	20	5.166		
7	408.6	5734	5	20	5.166		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-103018.mtw
 Equation: $Q = 0.0268037 \cdot A + 2.00798$
 RSD: 3.974 %
 Correlation coefficient: 0.999355




K3 = 0 K2 = 0 K1 = 0.0268037 K0 = 2.00798
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	6.894	105.4	0.15	20	9.024	
3	26.94	400.1	0.6	20	9.024	
4	54.35	805.7	1.2	20	9.024	
5	140.3	2071	3	20	9.024	
6	297	4310	6	20	9.024	
7	390.8	5625	7.5	20	9.024	

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-103018.mtw
 Equation: $Q = 0.0112873 \cdot A + 3.09951$
 RSD: 4.579 %
 Correlation coefficient: 0.999144




K3 = 0 K2 = 0 K1 = 0.0112873 K0 = 3.09951
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	11.76	193.9	0.15	20	11.64		
3	50.72	856.5	0.6	20	11.64		
4	105.8	1808	1.2	20	11.64		
5	276.9	4777	3	20	11.64		
6	590.6	1.012e+04	6	20	11.64		
7	772.2	1.329e+04	7.5	20	11.64		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-103018.mtw
 Equation: $Q = 0.0608473 \cdot A + 6.95891$
 RSD: 3.750 %
 Correlation coefficient: 0.999426




K3 = 0 K2 = 0 K1 = 0.0608473 K0 = 6.95891
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	5.643	140.6	0.5	20	15.43		
3	23.17	582.1	2	20	15.43		
4	46.9	1189	4	20	15.43		
5	119.7	3034	10	20	15.43		
6	249.3	6338	20	20	15.43		
7	324	8245	25	20	15.43		

1. ☐

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-103018.mtw
 Equation: $Q = 0.0109179 \cdot A + 1.79855$
 RSD: 3.799 %
 Correlation coefficient: 0.999411




K3 = 0 K2 = 0 K1 = 0.0109179 K0 = 1.79855
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	6.589	198.3	0.125	20	18.5		
3	26.55	806.6	0.5	20	18.5		
4	53.07	1644	1	20	18.5		
5	133.7	4217	2.5	20	18.5		
6	273.9	8828	5	20	18.5		
7	352.9	1.148e+04	6.25	20	18.5		

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-103018.mtw
 Equation: $Q = 0.0266646 \cdot A - 7.41229$
 RSD: 3.490 %
 Correlation coefficient: 0.999503




K3 = 0 K2 = 0 K1 = 0.0266646 K0 = -7.41229
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.276	306.2	0.25	20	23.19		
3	27.23	1099	1	20	23.19		
4	47.63	1908	2	20	23.19		
5	103.8	4058	5	20	23.19		
6	194.8	7640	10	20	23.19		
7	246.4	9722	12.5	20	23.19		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-103018.mtw
 Equation: $Q = 0.0351678 \cdot A + 14.2255$
 RSD: 4.849 %
 Correlation coefficient: 0.999040



K3 = 0 K2 = 0 K1 = 0.0351678 K0 = 14.2255
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	10.74	369.6	0.75	20	25.22		
3	41.42	1438	3	20	25.22		
4	83.34	2928	6	20	25.22		
5	215.8	7638	15	20	25.22		
6	455.8	1.629e+04	30	20	25.22		
7	592.8	2.137e+04	37.5	20	25.22		

Report date: 10/31/2018 11:28:00 AM
Printed by: wet


Ident: STD1
Analysis from: 10/30/2018 11:26:12 AM
File: _2018-10-30_

Last save: 10/31/2018 11:27:59 AM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18609

Last save: 10/30/2018 10:5

SAMPLE:
: AK/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/31/2018 11:28:12 AM
Printed by: wet


Ident: STD2
Analysis from: 10/30/2018 11:57:08 AM
File: _2018-10-30_

Last save: 10/30/2018 3:11:38 PM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18610

Last save: 10/30/2018 10:5

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.17	0.178	5.89	10.75	76.108	5.48
2	9.02	0.221	6.89	12.58	105.360	7.58
3	11.64	0.262	11.76	21.46	193.855	13.95
4	15.43	0.384	5.64	10.30	140.592	10.12
5	18.50	0.464	6.59	12.03	198.263	14.26
6	23.19	0.623	7.27	13.27	306.154	22.03
7	25.22	0.537	10.74	19.60	369.586	26.59
7	28.00	0.381	54.79	99.99	1389.917	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/31/2018 11:28:22 AM
Printed by: wet


Ident: STD3
Analysis from: 10/30/2018 12:28:05 PM
File: _2018-10-30_

Last save: 10/30/2018 3:11:38 PM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18611

Last save: 10/30/2018 10:5

SAMPLE:
: AK/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.17	0.181	28.61	12.73	408.308	7.30
2	9.02	0.221	26.94	11.99	400.059	7.16
3	11.63	0.265	50.72	22.58	856.472	15.32
4	15.38	0.383	23.17	10.32	582.115	10.41
5	18.42	0.465	26.55	11.82	806.642	14.43
6	23.19	0.602	27.23	12.12	1098.809	19.65
7	25.26	0.539	41.43	18.44	1438.460	25.73
7	28.00	0.379	224.64	100.00	5590.865	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/31/2018 11:28:32 AM
Printed by: wet


Ident: STD4
Analysis from: 10/30/2018 12:59:01 PM
File: _2018-10-30_

Last save: 10/30/2018 3:11:39 PM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18612

Last save: 10/30/2018 10:5

SAMPLE:
: AK/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.181	56.41	12.61	820.876	7.39
2	9.01	0.223	54.35	12.15	805.705	7.26
3	11.61	0.267	105.76	23.64	1808.343	16.29
4	15.33	0.387	46.90	10.48	1188.807	10.71
5	18.34	0.475	53.07	11.86	1643.928	14.81
6	23.16	0.598	47.63	10.64	1907.663	17.18
7	25.27	0.545	83.34	18.63	2928.026	26.37
7	28.00	0.382	447.45	100.00	11103.348	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/31/2018 11:28:44 AM
Printed by: wet


Ident: STD5
Analysis from: 10/30/2018 1:29:57 PM
File: _2018-10-30_

Last save: 10/30/2018 3:11:39 PM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18613

Last save: 10/30/2018 10:5

SAMPLE:
: AK/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.177	148.42	13.04	2064.408	7.41
2	9.00	0.221	140.25	12.32	2071.376	7.43
3	11.61	0.267	276.89	24.32	4777.004	17.15
4	15.25	0.388	119.72	10.51	3034.298	10.89
5	18.22	0.483	133.72	11.74	4217.393	15.14
6	23.12	0.588	103.82	9.12	4057.669	14.56
7	25.25	0.547	215.83	18.95	7638.161	27.42
7	28.00	0.382	1138.64	100.00	27860.308	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/31/2018 11:28:54 AM
Printed by: wet


Ident: STD6
Analysis from: 10/30/2018 2:00:53 PM
File: _2018-10-30_

Last save: 10/30/2018 3:11:39 PM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18614

Last save: 10/30/2018 10:5

SAMPLE:
: AK/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.19	0.185	318.19	13.37	4454.056	7.68
2	9.00	0.218	297.03	12.48	4310.136	7.43
3	11.62	0.264	590.60	24.82	10119.378	17.45
4	15.20	0.388	249.27	10.47	6337.912	10.93
5	18.12	0.492	273.91	11.51	8828.282	15.23
6	23.09	0.594	194.82	8.19	7639.807	13.18
7	25.25	0.552	455.84	19.16	16290.825	28.10
7	28.00	0.385	2379.66	100.00	57980.396	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/31/2018 11:29:05 AM
Printed by: wet


Ident: STD7
Analysis from: 10/30/2018 2:31:49 PM
File: _2018-10-30_

Last save: 10/30/2018 3:11:39 PM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18615

Last save: 10/30/2018 10:5

SAMPLE:
: AK/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.20	0.184	408.64	13.23	5734.363	7.60
2	8.99	0.217	390.76	12.66	5624.644	7.45
3	11.61	0.266	772.21	25.01	13289.056	17.61
4	15.15	0.389	323.98	10.49	8244.622	10.92
5	18.05	0.497	352.94	11.43	11481.834	15.21
6	23.06	0.599	246.43	7.98	9722.059	12.88
7	25.24	0.557	592.78	19.20	21370.681	28.32
7	28.00	0.387	3087.74	100.00	75467.259	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/31/2018 11:32:56 AM
Printed by: wet


Ident: ICV
Analysis from: 10/30/2018 3:14:58 PM
File: _2018-10-30_

Last save: 10/31/2018 11:32:46 AM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18616

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.19	0.179	148.19	13.08	2062.458	7.38
2	9.01	0.223	139.55	12.32	2066.654	7.40
3	11.60	0.269	276.79	24.44	4794.245	17.16
4	15.23	0.392	119.03	10.51	3036.950	10.87
5	18.18	0.489	132.76	11.72	4218.479	15.09
6	23.17	0.599	103.60	9.15	4122.550	14.75
7	25.34	0.556	212.72	18.78	7644.972	27.36
7	28.00	0.387	1132.63	100.00	27946.308	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/31/2018 11:33:14 AM
Printed by: wet


Ident: ICB
Analysis from: 10/30/2018 3:45:53 PM
File: _2018-10-30_

Last save: 10/31/2018 11:33:13 AM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18617

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 10
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 11:13:57 AM
Printed by: wet

IC-2 9056 A
water
LOD/LOQ


Ident: CCV
Analysis from: 10/30/2018 10:28:10 PM
File: _2018-10-30_

Last save: 10/31/2018 2:22:34 PM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18630

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.19	0.191	152.25	13.54	2243.775	7.81
2	8.99	0.236	135.39	12.04	2094.235	7.29
3	11.59	0.274	274.85	24.45	4851.527	16.88
4	15.21	0.418	114.24	10.16	3085.252	10.73
5	18.16	0.528	125.66	11.18	4293.677	14.94
6	23.10	0.616	108.65	9.66	4348.949	15.13
7	25.27	0.564	209.73	18.65	7752.526	26.97
7	28.00	0.404	1120.78	99.69	28669.940	99.75

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 11:14:06 AM
Printed by: wet


Ident: CCB
Analysis from: 10/30/2018 10:59:05 PM
File: _2018-10-30_

Last save: 10/31/2018 2:22:53 PM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18631

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 11:18:30 AM
Printed by: wet


Ident: LB98969BLW
Analysis from: 10/30/2018 11:30:03 PM
File: _2018-10-30_

Last save: 11/1/2018 11:18:27 AM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18632

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 13
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 12:20:57 PM
Printed by: wet


Ident: LB98969BSW
Analysis from: 10/31/2018 12:00:59 AM
File: _2018-10-31_

Last save: 11/1/2018 12:20:55 PM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18633

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.19	0.196	157.00	13.71	2365.669	8.02
2	9.00	0.240	136.70	11.94	2151.090	7.29
3	11.61	0.277	279.46	24.41	4985.268	16.89
4	15.24	0.426	115.17	10.06	3164.143	10.72
5	18.20	0.540	126.44	11.04	4406.959	14.93
6	23.09	0.611	112.15	9.80	4403.169	14.92
7	25.25	0.562	214.41	18.73	7960.996	26.98
7	28.00	0.407	1141.34	99.69	29437.294	99.75

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 11:19:19 AM
Printed by: wet


Ident: J5164-07
Analysis from: 10/31/2018 12:31:55 AM
File: _2018-10-31_

Last save: 10/31/2018 12:59:55 AM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18634

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.209	2.91	9.84	47.018	5.50
2	9.01	0.239	3.43	11.57	51.869	6.07
3	11.61	0.280	5.80	19.58	101.347	11.86
4	15.38	0.421	2.77	9.36	74.772	8.75
5	18.43	0.505	3.40	11.48	109.866	12.86
6	23.20	0.694	5.69	19.20	262.343	30.70
7	25.27	0.561	5.61	18.95	207.318	24.26
7	28.00	0.415	29.61	99.98	854.532	100.00

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 11:19:30 AM
Printed by: wet


Ident: J5164-09
Analysis from: 10/31/2018 1:02:50 AM
File: _2018-10-31_

Last save: 10/31/2018 1:30:51 AM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18635

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.210	3.27	10.95	53.951	6.26
2	9.02	0.246	3.46	11.56	55.836	6.48
3	11.62	0.278	5.72	19.11	99.154	11.50
4	15.40	0.424	2.80	9.36	77.863	9.03
5	18.46	0.506	3.37	11.25	107.946	12.52
6	23.20	0.690	5.73	19.17	262.766	30.48
7	25.28	0.558	5.56	18.58	204.599	23.73
7	28.00	0.416	29.91	99.98	862.115	100.00

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 11:20:22 AM
Printed by: wet


Ident: J5164-07
Analysis from: 10/31/2018 1:33:46 AM
File: _2018-10-31_

Last save: 11/1/2018 11:20:11 AM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18636

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.211	6.26	11.03	115.769	7.16
2	9.02	0.247	6.59	11.62	108.531	6.71
3	11.62	0.283	11.83	20.87	209.891	12.99
4	15.39	0.425	5.53	9.75	151.168	9.35
5	18.44	0.512	6.34	11.18	209.854	12.98
6	23.17	0.671	9.88	17.42	441.908	27.34
7	25.26	0.557	10.28	18.13	379.228	23.46
7	28.00	0.415	56.70	100.00	1616.348	100.00

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 11:20:34 AM
Printed by: wet


Ident: J5164-09
Analysis from: 10/31/2018 2:04:41 AM
File: _2018-10-31_

Last save: 11/1/2018 11:20:11 AM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18637

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 18
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.211	6.00	10.73	104.283	6.57
2	9.01	0.243	6.39	11.44	99.311	6.26
3	11.62	0.287	11.74	21.00	211.852	13.35
4	15.39	0.426	5.43	9.70	147.545	9.30
5	18.44	0.512	6.27	11.22	204.023	12.86
6	23.16	0.672	9.86	17.64	442.556	27.89
7	25.24	0.557	10.21	18.26	377.409	23.78
7	28.00	0.415	55.91	99.99	1586.979	100.00

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 11:20:53 AM
Printed by: wet


Ident: J5164-08
Analysis from: 10/31/2018 2:35:37 AM
File: _2018-10-31_

Last save: 10/31/2018 3:03:37 AM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18638

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 19
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.205	14.39	12.45	228.218	7.11
2	9.02	0.248	13.19	11.41	215.310	6.71
3	11.62	0.287	24.87	21.52	452.108	14.08
4	15.39	0.429	11.29	9.76	314.269	9.79
5	18.43	0.520	12.96	11.22	434.834	13.54
6	23.15	0.658	18.18	15.73	800.244	24.92
7	25.24	0.554	20.70	17.91	766.048	23.86
7	28.00	0.414	115.58	99.99	3211.031	100.00

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 12:22:04 PM
Printed by: wet


Ident: CCV
Analysis from: 10/31/2018 3:37:35 AM
File: _2018-10-31_

Last save: 11/1/2018 12:19:09 PM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18640

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.18	0.197	147.84	13.49	2256.243	7.81
2	9.01	0.245	131.16	11.97	2114.416	7.32
3	11.63	0.283	266.02	24.27	4882.058	16.91
4	15.28	0.438	109.95	10.03	3095.732	10.72
5	18.25	0.558	120.12	10.96	4305.768	14.91
6	23.04	0.608	108.97	9.94	4377.325	15.16
7	25.17	0.552	208.69	19.04	7775.435	26.93
7	28.00	0.412	1092.74	99.69	28806.978	99.76

This report has been created by IC Net
METROHM LTD

Report date: 11/1/2018 11:23:26 AM
Printed by: wet


Ident: CCB
Analysis from: 10/31/2018 4:08:35 AM
File: _2018-10-31_

Last save: 11/1/2018 11:23:25 AM

Method: AnionsIC2-103018.mtw
Run operator: wet
Analysis number: 18641

Last save: 10/30/2018 3:11

SAMPLE:
: AK/AP
Vial number: 21
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD