

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AR

11/2/2018 15:26

Test: ~~Total~~ CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.112	0.0	0.100	
ICB1	0.348	0.0	0.004	
CCV1	239.481	0.0	0.246	
CCB1	0.340	0.0	0.004	
PB114448BL	0.195	0.0	0.004	
J5729-04	0.023	0.0	0.004	
J5729-04DUP	0.132	0.0	0.004	
J5755-14	0.120	0.0	0.004	
J5778-07	0.367	0.0	0.004	
CCV2	242.410	0.0	0.249	
CCB2	0.615	0.0	0.004	

N	11
Mean	52.649
SD	97.2982
CV%	184.80

Aquakem v. 7.2AQ1

Results from time period:

Fri Nov 02 15:26:10 2018

Fri Nov 02 15:26:20 2018

AE

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A		Reactive Cl P	-0.195	µg/l	11/2/2018 9:59:23	
5.OPPBCN	A		Reactive Cl P	4.9469	µg/l	11/2/2018 9:59:24	
10PPBCN	A		Reactive Cl P	9.9673	µg/l	11/2/2018 9:59:25	
50PPBCN	A		Reactive Cl P	49.9108	µg/l	11/2/2018 9:59:26	
100PPBCN	A		Reactive Cl P	100.2024	µg/l	11/2/2018 9:59:27	
250PPBCN	A		Reactive Cl P	250.396	µg/l	11/2/2018 9:59:28	
500PPBCN	A		Reactive Cl P	499.7716	µg/l	11/2/2018 9:59:29	
ICV1	S		Reactive Cl P	95.1116	µg/l	11/2/2018 15:26:10	
ICB1	S		Reactive Cl P	0.3479	µg/l	11/2/2018 15:26:11	
CCV1	S		Reactive Cl P	239.4806	µg/l	11/2/2018 15:26:12	
CCB1	S		Reactive Cl P	0.3401	µg/l	11/2/2018 15:26:13	
PB114448BL	S		Reactive Cl P	0.1946	µg/l	11/2/2018 15:26:14	
J5729-04	S		Reactive Cl P	0.0226	µg/l	11/2/2018 15:26:15	
J5729-04DUP	S		Reactive Cl P	0.1324	µg/l	11/2/2018 15:26:16	
J5755-14	S		Reactive Cl P	0.1201	µg/l	11/2/2018 15:26:17	
J5778-07	S		Reactive Cl P	0.3671	µg/l	11/2/2018 15:26:18	
CCV2	S		Reactive Cl P	242.4103	µg/l	11/2/2018 15:26:19	
CCB2	S		Reactive Cl P	0.6148	µg/l	11/2/2018 15:26:20	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

11/2/2018 10:26


Test Total CN

Accepted 11/2/2018 10:26

Factor 989.3
Bias 0.004

Coeff. of det. 0.999999

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.1950	0.0000	
2	5.0PPBCN	0.009	4.9469	5.0000	
3	10PPBCN	0.014	9.9673	10.0000	
4	50PPBCN	0.054	49.9108	50.0000	
5	100PPBCN	0.105	100.2024	100.0000	
6	250PPBCN	0.257	250.3960	250.0000	
7	500PPBCN	0.509	499.7716	500.0000	