

689905 J

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AL

11/6/2018 11:14

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	92.257	0.0	0.098	
ICB1	-0.022	0.0	0.004	
CCV1	248.584	0.0	0.257	
CCB1	-0.149	0.0	0.004	
PB114480BL	-0.135	0.0	0.004	
PB114480BS	95.005	0.0	0.101	
MIDPB114480	251.253	0.0	0.260	
J5818-02	0.175	0.0	0.004	
J5818-02DUP	-0.028	0.0	0.004	
J5818-02MS	3.026	0.0	0.007	
J5818-02MSD	2.800	0.0	0.007	
CCV2	259.344	0.0	0.268	
CCB2	-0.352	0.0	0.003	

N	13
Mean	73.212
SD	108.0338
CV%	147.56

Aquakem v. 7.2AQ1

Results from time period:

Tue Nov 06 10:21:07 2018

Tue Nov 06 11:13:11 2018

Al

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.3484	µg/l	11/6/2018 10:40:43	
5.OPPBCN	A	Total CN	P	4.9163	µg/l	11/6/2018 10:40:44	
10PPBCN	A	Total CN	P	9.9339	µg/l	11/6/2018 10:40:45	
50PPBCN	A	Total CN	P	49.3515	µg/l	11/6/2018 10:40:46	
100PPBCN	A	Total CN	P	98.3438	µg/l	11/6/2018 10:40:47	
250PPBCN	A	Total CN	P	254.8093	µg/l	11/6/2018 10:40:48	
500PPBCN	A	Total CN	P	497.9937	µg/l	11/6/2018 10:40:49	
ICV1	S	Total CN	P	92.257	µg/l	11/6/2018 11:09:47	
ICB1	S	Total CN	P	-0.0224	µg/l	11/6/2018 11:09:48	
CCV1	S	Total CN	P	248.5835	µg/l	11/6/2018 11:09:49	
CCB1	S	Total CN	P	-0.1493	µg/l	11/6/2018 11:09:50	
PB114480BL	S	Total CN	P	-0.1351	µg/l	11/6/2018 11:09:51	
PB114480BS	S	Total CN	P	95.0047	µg/l	11/6/2018 11:09:52	
MIDPB114480	S	Total CN	P	251.2528	µg/l	11/6/2018 11:09:53	
J5818-02	S	Total CN	P	0.1749	µg/l	11/6/2018 11:09:54	
J5818-02DUP	S	Total CN	P	-0.0284	µg/l	11/6/2018 11:09:55	
J5818-02MS	S	Total CN	P	3.0257	µg/l	11/6/2018 11:09:56	
J5818-02MSD	S	Total CN	P	2.7998	µg/l	11/6/2018 11:09:57	
CCV2	S	Total CN	P	259.3443	µg/l	11/6/2018 11:13:10	
CCB2	S	Total CN	P	-0.3515	µg/l	11/6/2018 11:13:11	

LB99051
CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AR

11/6/2018 10:41

Test Total CN


Accepted 11/6/2018 10:41

Factor 980.9

Bias 0.004

Coeff. of det. 0.999852

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.3484	0.0000	
2	5.0PPBCN	0.009	4.9163	5.0000	
3	10PPBCN	0.014	9.9339	10.0000	
4	50PPBCN	0.054	49.3515	50.0000	
5	100PPBCN	0.104	98.3438	100.0000	
6	250PPBCN	0.264	254.8093	250.0000	
7	500PPBCN	0.511	497.9937	500.0000	