

XB99161

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/12/2018 13:32

Test: Total CN µg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.280	0.0	0.097	
ICB1	-0.608	0.0	0.004	
CCV1	243.634	0.0	0.235	
CCB1	-0.914	0.0	0.004	
PB114724BL	-1.149	0.0	0.003	
J5766-02	-1.195	0.0	0.003	
J5766-02DUP	-1.457	0.0	0.003	
J5892-02	-0.982	0.0	0.004	
J5901-08	-1.525	0.0	0.003	
J5907-04	-1.110	0.0	0.003	
J5907-08	-1.003	0.0	0.004	
J5907-12	-1.014	0.0	0.004	
J5907-16	-1.072	0.0	0.003	
CCV2	257.283	0.0	0.248	
CCB2	-0.661	0.0	0.004	

N	15
Mean	39.034
SD	89.5126
CV%	229.32

Aquakem v. 7.2AQ1

Results from time period:

Mon Nov 12 12:59:37 2018

Mon Nov 12 13:32:00 2018

AB

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.9228	µg/l	11/12/2018 12:59:37	
5.OPPBCN	A	Reactive CN	P	4.077	µg/l	11/12/2018 12:59:38	
10PPBCN	A	Reactive CN	P	10.9218	µg/l	11/12/2018 12:59:39	
50PPBCN	A	Reactive CN	P	49.1377	µg/l	11/12/2018 12:59:40	
100PPBCN	A	Reactive CN	P	101.5177	µg/l	11/12/2018 12:59:41	
250PPBCN	A	Reactive CN	P	250.9902	µg/l	11/12/2018 12:59:42	
500PPBCN	A	Reactive CN	P	499.2784	µg/l	11/12/2018 12:59:43	
ICV1	S	Reactive CN	P	97.2796	µg/l	11/12/2018 13:27:38	
ICB1	S	Reactive CN	P	-0.608	µg/l	11/12/2018 13:27:39	
CCV1	S	Reactive CN	P	243.6344	µg/l	11/12/2018 13:27:40	
CCB1	S	Reactive CN	P	-0.9144	µg/l	11/12/2018 13:27:41	
PB114724BL	S	Reactive CN	P	-1.1487	µg/l	11/12/2018 13:27:42	
J5766-02	S	Reactive CN	P	-1.1955	µg/l	11/12/2018 13:27:43	
J5766-02DUP	S	Reactive CN	P	-1.4568	µg/l	11/12/2018 13:27:44	
J5892-02	S	Reactive CN	P	-0.9822	µg/l	11/12/2018 13:27:45	
J5901-08	S	Reactive CN	P	-1.5248	µg/l	11/12/2018 13:27:46	
J5907-04	S	Reactive CN	P	-1.11	µg/l	11/12/2018 13:27:47	
J5907-08	S	Reactive CN	P	-1.0033	µg/l	11/12/2018 13:27:48	
J5907-12	S	Reactive CN	P	-1.0136	µg/l	11/12/2018 13:31:57	
J5907-16	S	Reactive CN	P	-1.0718	µg/l	11/12/2018 13:31:58	
CCV2	S	Reactive CN	P	257.2828	µg/l	11/12/2018 13:31:59	
CCB2	S	Reactive CN	P	-0.6612	µg/l	11/12/2018 13:32:00	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/12/2018 13:00


Test Total CN

Accepted 11/12/2018 13:00

Factor 1055
Bias 0.004

Coeff. of det. 0.999965

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.9228	0.0000	
2	5.0PPBCN	0.008	4.0770	5.0000	
3	10PPBCN	0.015	10.9218	10.0000	
4	50PPBCN	0.051	49.1377	50.0000	
5	100PPBCN	0.101	101.5177	100.0000	
6	250PPBCN	0.242	250.9902	250.0000	
7	500PPBCN	0.478	499.2784	500.0000	