

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

11/15/2018 13:52

Test: ~~Total~~ ^{Reactive} CN ^{ug/L}

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.069	0.0	0.087	
ICB1	0.313	0.0	0.004	
CCV1	240.105	0.0	0.212	
CCB1	0.120	0.0	0.003	
PB114849BL	-0.065	0.0	0.003	
J5767-10	0.210	0.0	0.003	
J5767-10DUP	-0.222	0.0	0.003	
J5767-12	-0.133	0.0	0.003	
J5767-14	0.073	0.0	0.003	
J5767-16	-0.079	0.0	0.003	
J5942-04	-0.156	0.0	0.003	
J5942-08	-0.175	0.0	0.003	
J5942-12	-0.322	0.0	0.003	
J5954-01	-0.152	0.0	0.003	
CCV2	250.314	0.0	0.221	
CCB2	0.271	0.0	0.004	

N	16
Mean	36.636
SD	84.8773
CV%	231.68

Aquakem v. 7.2AQ1

Results from time period:

Thu Nov 15 13:47:20 2018

Thu Nov 15 13:52:11 2018

AS

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.184	µg/l	11/15/2018 9:55:10	
5.OPPBCN	A	Reactive CN	P	4.63	µg/l	11/15/2018 9:55:11	
10PPBCN	A	Reactive CN	P	10.3343	µg/l	11/15/2018 9:55:12	
50PPBCN	A	Reactive CN	P	49.9577	µg/l	11/15/2018 9:55:13	
100PPBCN	A	Reactive CN	P	99.8643	µg/l	11/15/2018 9:55:14	
250PPBCN	A	Reactive CN	P	250.7386	µg/l	11/15/2018 9:55:15	
500PPBCN	A	Reactive CN	P	499.6591	µg/l	11/15/2018 9:55:16	
ICV1	S	Reactive CN	P	96.069	µg/l	11/15/2018 13:47:20	
ICB1	S	Reactive CN	P	0.3133	µg/l	11/15/2018 13:47:21	
CCV1	S	Reactive CN	P	240.1052	µg/l	11/15/2018 13:47:22	
CCB1	S	Reactive CN	P	0.1199	µg/l	11/15/2018 13:47:23	
PB114849BL	S	Reactive CN	P	-0.0648	µg/l	11/15/2018 13:47:24	
J5767-10	S	Reactive CN	P	0.2104	µg/l	11/15/2018 13:47:25	
J5767-10DUP	S	Reactive CN	P	-0.2217	µg/l	11/15/2018 13:47:26	
J5767-12	S	Reactive CN	P	-0.1335	µg/l	11/15/2018 13:47:27	
J5767-14	S	Reactive CN	P	0.0727	µg/l	11/15/2018 13:47:28	
J5767-16	S	Reactive CN	P	-0.0789	µg/l	11/15/2018 13:47:29	
J5942-04	S	Reactive CN	P	-0.156	µg/l	11/15/2018 13:47:30	
J5942-08	S	Reactive CN	P	-0.1754	µg/l	11/15/2018 13:52:07	
J5942-12	S	Reactive CN	P	-0.3222	µg/l	11/15/2018 13:52:08	
J5954-01	S	Reactive CN	P	-0.1519	µg/l	11/15/2018 13:52:09	
CCV2	S	Reactive CN	P	250.3135	µg/l	11/15/2018 13:52:10	
CCB2	S	Reactive CN	P	0.2707	µg/l	11/15/2018 13:52:11	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/15/2018 9:55


Test Total CN

Accepted 11/15/2018 9:55

Factor 1150
Bias 0.003

Coeff. of det. 0.999995

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.1840	0.0000	
2	5.0PPBCN	0.007	4.6300	5.0000	
3	10PPBCN	0.012	10.3343	10.0000	
4	50PPBCN	0.047	49.9577	50.0000	
5	100PPBCN	0.090	99.8643	100.0000	
6	250PPBCN	0.221	250.7386	250.0000	
7	500PPBCN	0.438	499.6591	500.0000	