

LB 99258


7471B

INSTRUMENT ID : CV1

Linear

 μ Abs.:

55202



A= 0.0000e+000

B= 1.8183e-004

C= 1.8789e-003

Rho= 0.9999660

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	61	0.000	61				
0.2	0.200	0.190	-0.010	1036	0.0 %	1036				
2.5	2.500	2.514	0.014	13818	0.0 %	13818				
5.0	5.000	5.002	0.002	27500	0.0 %	27500				
7.5	7.500	7.441	-0.059	40910	0.0 %	40910				
10.0	10.000	10.039	0.039	55202	0.0 %	55202				

LB99258 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Nov 2018 16:59:18

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	61	-	0.00007471B	PPB	15 Nov 2018 17:01:30	S	Std
0.2 - 1	50.2	1036	-	0.20007471B	PPB	15 Nov 2018 17:03:48	S	Std
2.5 - 1	52.5	13818	-	2.50007471B	PPB	15 Nov 2018 17:06:06	S	Std
5.0 - 1	55	27500	-	5.00007471B	PPB	15 Nov 2018 17:08:23	S	Std
7.5 - 1	57.5	40910	-	7.50007471B	PPB	15 Nov 2018 17:10:41	S	Std
10.0 - 1	59.0	55202	-	10.00007471B	PPB	15 Nov 2018 17:12:59	S	Std
ICV77 - 1	ICV77	21235	3.8631	-7471B	PPB	15 Nov 2018 17:15:47	U	SMPL
ICB77 - 1	ICB77	-69	-0.0107	-7471B	PPB	15 Nov 2018 17:18:03	U	SMPL
CCV24 - 1	CCV24	27503	5.0028	-7471B	PPB	15 Nov 2018 17:20:20	U	SMPL
CCB24 - 1	CCB24	-77	-0.0121	-7471B	PPB	15 Nov 2018 17:22:32	U	SMPL
CRA - 1	CRA	1002	0.1841	-7471B	PPB	15 Nov 2018 17:24:53	U	SMPL
HighStd - 1	HighStd	54961	9.9956	-7471B	PPB	15 Nov 2018 17:27:09	U	SMPL
ChkStd - 1	ChkStd	36781	6.6899	-7471B	PPB	15 Nov 2018 17:29:25	U	SMPL
PB114833BL - 1	PBS	-117	-0.0194	-7471B	PPB	15 Nov 2018 17:31:44	U	SMPL
PB114833BS - 1	LCSS	17742	3.2280	-7471B	PPB	15 Nov 2018 17:34:03	U	SMPL
J5767-09 - 1	CL-01-111418-A	7636	1.3904	-7471B	PPB	15 Nov 2018 17:36:21	U	SMPL
J5767-11 - 1	CL-01-111418-B	3636	0.6630	-7471B	PPB	15 Nov 2018 17:38:39	U	SMPL
J5767-13 - 1	CL-01-111418-C	3727	0.6796	-7471B	PPB	15 Nov 2018 17:40:56	U	SMPL
J5767-15 - 1	CL-01-111418-D	2654	0.4845	-7471B	PPB	15 Nov 2018 17:43:14	U	SMPL
J5942-01 - 1	VW-D1	34838	6.3366	-7471B	PPB	15 Nov 2018 17:45:31	U	SMPL
CCV25 - 1	CCV25	27221	4.9515	-7471B	PPB	15 Nov 2018 17:47:47	U	SMPL
CCB25 - 1	CCB25	-54	-0.0079	-7471B	PPB	15 Nov 2018 17:50:05	U	SMPL
J5942-05 - 1	VW-F1	376	0.0702	-7471B	PPB	15 Nov 2018 17:52:30	U	SMPL
J5942-09 - 1	VW-F2	447	0.0832	-7471B	PPB	15 Nov 2018 17:54:47	U	SMPL
J5950-02 - 1	SSI-04	1511	0.2766	-7471B	PPB	15 Nov 2018 17:57:03	U	SMPL
J5950-03 - 1	SSI-03	2627	0.4796	-7471B	PPB	15 Nov 2018 17:59:19	U	SMPL
J5950-04 - 1	SSI-02	655	0.1210	-7471B	PPB	15 Nov 2018 18:01:36	U	SMPL
J5950-05 - 1	SSI-01	1813	0.3315	-7471B	PPB	15 Nov 2018 18:03:53	U	SMPL
J5950-06 - 1	SSI-20-TW-1	2453	0.4479	-7471B	PPB	15 Nov 2018 18:06:10	U	SMPL
J5950-07 - 1	SSI-21-TW-2	777	0.1432	-7471B	PPB	15 Nov 2018 18:08:28	U	SMPL
J5950-07DUP - 1	SSI-21-TW-2DUP	801	0.1475	-7471B	PPB	15 Nov 2018 18:13:25	U	SMPL
J5950-08 - 1	J5950-07MS	19018	3.4600	-7471B	PPB	15 Nov 2018 18:15:41	U	SMPL
CCV26 - 1	CCV26	27372	4.9790	-7471B	PPB	15 Nov 2018 18:17:57	U	SMPL
CCB26 - 1	CCB26	-64	-0.0098	-7471B	PPB	15 Nov 2018 18:20:14	U	SMPL
J5950-09 - 1	J5950-07MSD	18688	3.4000	-7471B	PPB	15 Nov 2018 18:22:31	U	SMPL
J5950-10 - 1	SSI-FD-11132018	1089	0.1999	-7471B	PPB	15 Nov 2018 18:24:48	U	SMPL
J5950-12 - 1	SSI-15	282	0.0532	-7471B	PPB	15 Nov 2018 18:27:05	U	SMPL
J5950-14 - 1	SSI-08	1800	0.3292	-7471B	PPB	15 Nov 2018 18:29:22	U	SMPL
J5950-07LX5 - 1		205	0.0392	-7471B	PPB	15 Nov 2018 18:31:39	U	SMPL
J5950-07A - 1		23977	4.3617	-7471B	PPB	15 Nov 2018 18:33:56	U	SMPL
CCV27 - 1	CCV27	27572	5.0154	-7471B	PPB	15 Nov 2018 18:36:13	U	SMPL
CCB27 - 1	CCB27	-55	-0.0081	-7471B	PPB	15 Nov 2018 18:38:32	U	SMPL