

LB99302

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : ds

11/19/2018 14:07

Test: *Reactive*
~~Total~~ CN

Sample Id	<i>mg/L</i> Result	Dil. 1 +	Response	Errors
ICV1	95.218	0.0	0.090	
ICB1	-0.185	0.0	0.004	
CCV1	241.304	0.0	0.221	
CCB1	-0.352	0.0	0.004	
PB114914BL	-0.873	0.0	0.003	
J5774-02	-0.463	0.0	0.003	
J5774-02DUP	-0.565	0.0	0.003	
J5989-02	-0.506	0.0	0.003	
J5991-02	-0.374	0.0	0.004	
PB114915BL	-0.522	0.0	0.003	
J5980-02	-0.537	0.0	0.003	
J5980-02DUP	-0.352	0.0	0.004	
CCV2	258.254	0.0	0.236	
CCB2	-0.012	0.0	0.004	

N	14
Mean	42.145
SD	91.6199
CV%	217.39

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : ds

11/19/2018 14:54


Test Total CN

Accepted 11/19/2018 10:15

Factor 1112
Bias 0.004

Coeff. of det. 0.999941

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.8730	0.0000	
2	5.0PPBCN	0.008	4.2451	5.0000	
3	10PPBCN	0.012	9.3848	10.0000	
4	50PPBCN	0.049	49.7432	50.0000	
5	100PPBCN	0.095	101.5817	100.0000	
6	250PPBCN	0.231	252.3780	250.0000	
7	500PPBCN	0.452	498.5402	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Mon Nov 19 14:03:18 2018

Mon Nov 19 14:07:11 2018

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.873	µg/l	11/19/2018 10:05:10	
5.OPPBCN	A	Reactive CN	P	4.2451	µg/l	11/19/2018 10:05:11	
10PPBCN	A	Reactive CN	P	9.3848	µg/l	11/19/2018 10:05:12	
50PPBCN	A	Reactive CN	P	49.7432	µg/l	11/19/2018 10:05:13	
100PPBCN	A	Reactive CN	P	101.5817	µg/l	11/19/2018 10:05:14	
250PPBCN	A	Reactive CN	P	252.378	µg/l	11/19/2018 10:05:15	
500PPBCN	A	Reactive CN	P	498.5402	µg/l	11/19/2018 10:05:16	
ICV1	S	Reactive CN	P	95.2184	µg/l	11/19/2018 14:03:18	
ICB1	S	Reactive CN	P	-0.1848	µg/l	11/19/2018 14:03:19	
CCV1	S	Reactive CN	P	241.3039	µg/l	11/19/2018 14:03:20	
CCB1	S	Reactive CN	P	-0.3517	µg/l	11/19/2018 14:03:21	
PB114914BL	S	Reactive CN	P	-0.8735	µg/l	11/19/2018 14:03:22	
J5774-02	S	Reactive CN	P	-0.463	µg/l	11/19/2018 14:03:23	
J5774-02DUP	S	Reactive CN	P	-0.5646	µg/l	11/19/2018 14:03:24	
J5989-02	S	Reactive CN	P	-0.5057	µg/l	11/19/2018 14:03:25	
J5991-02	S	Reactive CN	P	-0.3743	µg/l	11/19/2018 14:03:26	
PB114915BL	S	Reactive CN	P	-0.5221	µg/l	11/19/2018 14:03:27	
J5980-02	S	Reactive CN	P	-0.537	µg/l	11/19/2018 14:03:28	
J5980-02DUP	S	Reactive CN	P	-0.3521	µg/l	11/19/2018 14:07:09	
CCV2	S	Reactive CN	P	258.2544	µg/l	11/19/2018 14:07:10	
CCB2	S	Reactive CN	P	-0.0116	µg/l	11/19/2018 14:07:11	