

21399150

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/12/2018 7:06

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.380	0.0	0.089	
ICB1	-0.963	0.0	0.004	
CCV1	238.923	0.0	0.215	
CCB1	-0.889	0.0	0.004	
PB115515BL	-1.056	0.0	0.004	
PB115515BS	96.330	0.0	0.089	
MIDPB115515	240.325	0.0	0.216	
J6312-02	62.779	0.0	0.060	
J6312-02DUP	62.279	0.0	0.059	
J6312-02MS	327.224	0.0	0.292	
J6312-02MSD	327.394	0.0	0.292	
CCV2	249.393	0.0	0.224	
CCB2	-1.492	0.0	0.003	

N	13
Mean	130.510
SD	127.8887
CV%	97.99

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/11/2018 15:47


Test Total CN

Accepted 12/11/2018 7:44

Factor 1138
Bias 0.005

Coeff. of det. 0.999894

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.4118	0.0000	
2	5.0PPBCN	0.008	3.9441	5.0000	
3	10PPBCN	0.015	11.9262	10.0000	
4	50PPBCN	0.047	47.6744	50.0000	
5	100PPBCN	0.094	102.1365	100.0000	
6	250PPBCN	0.226	251.9067	250.0000	
7	500PPBCN	0.443	498.8240	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Tue Dec 11 16:21:19 2018

Tue Dec 11 16:45:54 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-1.4118	µg/l	12/11/2018 7:43:54	
5.OPPBCN	A	Total CN	P	3.9441	µg/l	12/11/2018 7:43:55	
10PPBCN	A	Total CN	P	11.9262	µg/l	12/11/2018 7:43:56	
50PPBCN	A	Total CN	P	47.6744	µg/l	12/11/2018 7:43:57	
100PPBCN	A	Total CN	P	102.1365	µg/l	12/11/2018 7:43:58	
250PPBCN	A	Total CN	P	251.9067	µg/l	12/11/2018 7:43:59	
500PPBCN	A	Total CN	P	498.824	µg/l	12/11/2018 7:44:00	
ICV1	S	Total CN	P	96.3801	µg/l	12/11/2018 16:21:19	
ICB1	S	Total CN	P	-0.9628	µg/l	12/11/2018 16:21:20	
CCV1	S	Total CN	P	238.9226	µg/l	12/11/2018 16:21:21	
CCB1	S	Total CN	P	-0.8887	µg/l	12/11/2018 16:21:22	
PB115515BL	S	Total CN	P	-1.0559	µg/l	12/11/2018 16:21:23	
PB115515BS	S	Total CN	P	96.3296	µg/l	12/11/2018 16:21:24	
MIDPB115515	S	Total CN	P	240.3247	µg/l	12/11/2018 16:21:25	
J6312-02	S	Total CN	P	62.7787	µg/l	12/11/2018 16:45:49	
J6312-02DUP	S	Total CN	P	62.2792	µg/l	12/11/2018 16:45:50	
J6312-02MS	S	Total CN	P	327.2237	µg/l	12/11/2018 16:45:51	
J6312-02MSD	S	Total CN	P	327.3939	µg/l	12/11/2018 16:45:52	
CCV2	S	Total CN	P	249.3934	µg/l	12/11/2018 16:45:53	
CCB2	S	Total CN	P	-1.4919	µg/l	12/11/2018 16:45:54	