

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : As

12/18/2018 11:26

Test: ~~Total~~ ^{Reactive} CN

Sample Id	Result ^{ug/l}	Dil. 1 +	Response	Errors
ICV1	93.468	0.0	0.086	
ICB1	0.796	0.0	0.004	
CCV1	234.917	0.0	0.212	
CCB1	0.388	0.0	0.004	
PB115768BL	0.439	0.0	0.004	
J6195-10	0.639	0.0	0.004	
J6195-10DUP	0.165	0.0	0.003	
J6429-04	0.314	0.0	0.003	
J6453-03	0.714	0.0	0.004	
J6462-01	0.576	0.0	0.004	
J6464-03	0.680	0.0	0.004	
J6464-06	0.172	0.0	0.003	
PB115800BL	0.475	0.0	0.004	
J6462-07	0.478	0.0	0.004	
CCV2	239.023	0.0	0.216	
CCB2	0.288	0.0	0.003	
J6462-07DUP	0.544	0.0	0.004	
CCV3	248.252	0.0	0.224	
CCB3	0.746	0.0	0.004	

N	19
Mean	43.320
SD	90.3794
CV%	208.63

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

12/18/2018 9:39


Test Total CN

Accepted 12/18/2018 9:39

Factor 1125
Bias 0.003

Coeff. of det. 0.999958

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.5977	0.0000	
2	5.0PPBCN	0.007	4.7648	5.0000	
3	10PPBCN	0.012	9.8972	10.0000	
4	50PPBCN	0.046	47.9767	50.0000	
5	100PPBCN	0.093	100.9307	100.0000	
6	250PPBCN	0.227	251.6245	250.0000	
7	500PPBCN	0.447	499.2083	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Tue Dec 18 09:36:39 2018

Tue Dec 18 11:25:51 2018

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.5977	µg/l	12/18/2018 9:36:39	
5.OPPBCN	A	Reactive CN	P	4.7648	µg/l	12/18/2018 9:36:40	
10PPBCN	A	Reactive CN	P	9.8972	µg/l	12/18/2018 9:36:41	
50PPBCN	A	Reactive CN	P	47.9767	µg/l	12/18/2018 9:36:42	
100PPBCN	A	Reactive CN	P	100.9307	µg/l	12/18/2018 9:36:43	
250PPBCN	A	Reactive CN	P	251.6245	µg/l	12/18/2018 9:36:44	
500PPBCN	A	Reactive CN	P	499.2083	µg/l	12/18/2018 9:36:45	
ICV1	S	Reactive CN	P	93.4683	µg/l	12/18/2018 11:19:33	
ICB1	S	Reactive CN	P	0.7959	µg/l	12/18/2018 11:19:34	
CCV1	S	Reactive CN	P	234.9165	µg/l	12/18/2018 11:19:35	
CCB1	S	Reactive CN	P	0.3877	µg/l	12/18/2018 11:19:36	
PB115768BL	S	Reactive CN	P	0.4386	µg/l	12/18/2018 11:19:37	
J6195-10	S	Reactive CN	P	0.6389	µg/l	12/18/2018 11:19:38	
J6195-10DUP	S	Reactive CN	P	0.1649	µg/l	12/18/2018 11:19:39	
J6429-04	S	Reactive CN	P	0.3137	µg/l	12/18/2018 11:19:40	
J6453-03	S	Reactive CN	P	0.7143	µg/l	12/18/2018 11:19:41	
J6462-01	S	Reactive CN	P	0.5758	µg/l	12/18/2018 11:19:42	
J6464-03	S	Reactive CN	P	0.6801	µg/l	12/18/2018 11:19:43	
J6464-06	S	Reactive CN	P	0.1722	µg/l	12/18/2018 11:25:44	
PB115800BL	S	Reactive CN	P	0.4754	µg/l	12/18/2018 11:25:45	
J6462-07	S	Reactive CN	P	0.4783	µg/l	12/18/2018 11:25:46	
CCV2	S	Reactive CN	P	239.0225	µg/l	12/18/2018 11:25:47	
CCB2	S	Reactive CN	P	0.2876	µg/l	12/18/2018 11:25:48	
J6462-07DUP	S	Reactive CN	P	0.5444	µg/l	12/18/2018 11:25:49	
CCV3	S	Reactive CN	P	248.2517	µg/l	12/18/2018 11:25:50	
CCB3	S	Reactive CN	P	0.7462	µg/l	12/18/2018 11:25:51	