

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : ds

12/19/2018 12:12

Test: Residue
Total CN

Sample Id	Result <u>ug/L</u>	Dil. 1 +	Response	Errors
ICV1	95.136	0.0	0.088	
ICB1	-0.061	0.0	0.004	
CCV1	239.497	0.0	0.215	
CCB1	-0.302	0.0	0.004	
PB115846BL	-0.283	0.0	0.004	
J6194-08	-0.124	0.0	0.004	
J6194-08DUP	-0.499	0.0	0.003	
J6194-10	-0.363	0.0	0.004	
J6196-08	-0.374	0.0	0.004	
J6196-12	-0.187	0.0	0.004	
J6466-01	-0.298	0.0	0.004	
J6469-02	-0.655	0.0	0.003	
CCV2	240.066	0.0	0.216	
CCB2	-0.020	0.0	0.004	
J6475-02	-0.057	0.0	0.004	
PB115847BL	-0.291	0.0	0.004	
J6472-04	-0.445	0.0	0.003	
J6472-04DUP	-0.418	0.0	0.004	
J6199-10	-0.423	0.0	0.004	
J6199-12	-0.411	0.0	0.004	
J6194-10	-0.834	0.0	0.003	
J6472-01	0.525	0.0	0.004	
CCV3	253.219	0.0	0.227	
CCB3	-0.474	0.0	0.003	

N 24
Mean 34.247
SD 83.4099
CV% 243.55

J6199-12
J6196-10

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Dec 19 09:19:37 2018

Wed Dec 19 12:07:13 2018

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.2612	µg/l	12/19/2018 9:19:37	
5.OPPBCN	A	Reactive CN	P	4.5346	µg/l	12/19/2018 9:19:38	
10PPBCN	A	Reactive CN	P	9.42	µg/l	12/19/2018 9:19:39	
50PPBCN	A	Reactive CN	P	48.118	µg/l	12/19/2018 9:19:40	
100PPBCN	A	Reactive CN	P	101.7654	µg/l	12/19/2018 9:19:41	
250PPBCN	A	Reactive CN	P	253.1436	µg/l	12/19/2018 9:19:42	
500PPBCN	A	Reactive CN	P	498.2796	µg/l	12/19/2018 9:19:43	
ICV1	S	Reactive CN	P	95.1363	µg/l	12/19/2018 11:36:57	
ICB1	S	Reactive CN	P	-0.0607	µg/l	12/19/2018 11:36:58	
CCV1	S	Reactive CN	P	239.4965	µg/l	12/19/2018 11:36:59	
CCB1	S	Reactive CN	P	-0.3018	µg/l	12/19/2018 11:37:00	
PB115846BL	S	Reactive CN	P	-0.2831	µg/l	12/19/2018 11:37:01	
J6194-08	S	Reactive CN	P	-0.1239	µg/l	12/19/2018 11:37:02	
J6194-08DUP	S	Reactive CN	P	-0.4988	µg/l	12/19/2018 11:37:03	
J6194-10	S	Reactive CN	P	-0.3635	µg/l	12/19/2018 11:37:04	
J6196-08	S	Reactive CN	P	-0.3741	µg/l	12/19/2018 11:37:05	
J6196-12	S	Reactive CN	P	-0.1867	µg/l	12/19/2018 11:37:07	
J6466-01	S	Reactive CN	P	-0.2976	µg/l	12/19/2018 11:44:32	
J6469-02	S	Reactive CN	P	-0.6546	µg/l	12/19/2018 11:44:33	
CCV2	S	Reactive CN	P	240.066	µg/l	12/19/2018 11:44:35	
CCB2	S	Reactive CN	P	-0.0201	µg/l	12/19/2018 11:44:36	
J6475-02	S	Reactive CN	P	-0.057	µg/l	12/19/2018 11:44:37	
PB115847BL	S	Reactive CN	P	-0.2909	µg/l	12/19/2018 11:44:38	
J6472-04	S	Reactive CN	P	-0.4455	µg/l	12/19/2018 11:44:39	
J6472-04DUP	S	Reactive CN	P	-0.418	µg/l	12/19/2018 11:44:40	
J6199-10	S	Reactive CN	P	-0.423	µg/l	12/19/2018 11:44:41	
J6199-12	S	Reactive CN	P	-0.4109	µg/l	12/19/2018 11:44:42	
J6196-10	S	Reactive CN	P	-0.8336	µg/l	12/19/2018 12:07:10	
J6472-01	S	Reactive CN	P	0.5254	µg/l	12/19/2018 12:07:11	
CCV3	S	Reactive CN	P	253.2193	µg/l	12/19/2018 12:07:12	
CCB3	S	Reactive CN	P	-0.4738	µg/l	12/19/2018 12:07:13	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/19/2018 9:20


Test Total CN

Accepted 12/19/2018 9:20

Factor 1133
Bias 0.004

Coeff. of det. 0.999902

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.2612	0.0000	
2	5.0PPBCN	0.008	4.5346	5.0000	
3	10PPBCN	0.012	9.4200	10.0000	
4	50PPBCN	0.046	48.1180	50.0000	
5	100PPBCN	0.094	101.7654	100.0000	
6	250PPBCN	0.227	253.1436	250.0000	
7	500PPBCN	0.444	498.2796	500.0000	