

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : LS

12/20/2018 13:00

Test: *Acetone*
~~Total~~ CN

Sample Id	Result <i>Mg/L</i>	Dil. 1 +	Response	Errors
ICV1	98.037	0.0	0.080	
ICB1	0.431	0.0	0.004	
CCV1	245.929	0.0	0.194	
CCB1	0.014	0.0	0.004	
PB115884BL	-0.126	0.0	0.004	
J6482-04	-0.226	0.0	0.004	
J6482-04DUP	-0.084	0.0	0.004	
J6482-12	-0.014	0.0	0.004	
J6493-02	-0.179	0.0	0.004	
J6494-01	-0.148	0.0	0.004	
CCV2	228.220	0.0	0.181	
CCB2	-0.036	0.0	0.004	

N	12
Mean	47.652
SD	92.8973
CV%	194.95

As

Aquakem v. 7.2AQ1

Results from time period:

Thu Dec 20 11:58:59 2018

Thu Dec 20 12:59:40 2018

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.239	µg/l	12/20/2018 12:33:31	
5.0PPBCN	A	Reactive CN	P	5.5361	µg/l	12/20/2018 12:33:32	
10PPBCN	A	Reactive CN	P	10.0636	µg/l	12/20/2018 12:33:33	
50PPBCN	A	Reactive CN	P	50.2902	µg/l	12/20/2018 12:33:34	
100PPBCN	A	Reactive CN	P	101.1613	µg/l	12/20/2018 12:33:35	
250PPBCN	A	Reactive CN	P	246.9114	µg/l	12/20/2018 12:33:36	
500PPBCN	A	Reactive CN	P	501.2764	µg/l	12/20/2018 12:33:37	
ICV1	S	Reactive CN	P	98.037	µg/l	12/20/2018 12:56:52	
ICB1	S	Reactive CN	P	0.4305	µg/l	12/20/2018 12:56:53	
CCV1	S	Reactive CN	P	245.9295	µg/l	12/20/2018 12:56:54	
CCB1	S	Reactive CN	P	0.0141	µg/l	12/20/2018 12:56:55	
PB115884BL	S	Reactive CN	P	-0.1261	µg/l	12/20/2018 12:56:56	
J6482-04	S	Reactive CN	P	-0.2257	µg/l	12/20/2018 12:56:57	
J6482-04DUP	S	Reactive CN	P	-0.0843	µg/l	12/20/2018 12:56:58	
J6482-12	S	Reactive CN	P	-0.0139	µg/l	12/20/2018 12:56:59	
J6493-02	S	Reactive CN	P	-0.1786	µg/l	12/20/2018 12:57:00	
J6494-01	S	Reactive CN	P	-0.1478	µg/l	12/20/2018 12:57:01	
CCV2	S	Reactive CN	P	228.2196	µg/l	12/20/2018 12:57:02	
CCB2	S	Reactive CN	P	-0.0363	µg/l	12/20/2018 12:59:40	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/20/2018 12:35


Test Total CN

Accepted 12/20/2018 12:34

Factor 1291
Bias 0.004

Coeff. of det. 0.999937

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.2390	0.0000	
2	5.0PPBCN	0.008	5.5361	5.0000	
3	10PPBCN	0.012	10.0636	10.0000	
4	50PPBCN	0.043	50.2902	50.0000	
5	100PPBCN	0.082	101.1613	100.0000	
6	250PPBCN	0.195	246.9114	250.0000	
7	500PPBCN	0.392	501.2764	500.0000	