

=====

Aquakem 7.2AQ1

Page: 1

LB9999/

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

12/21/2018 12:20

Test: ~~Total~~ ^{Reactive} CNMg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.116	0.0	0.085	
ICB1	0.042	0.0	0.004	
CCV1	235.612	0.0	0.203	
CCB1	-0.367	0.0	0.004	
PB115923BL	-0.174	0.0	0.004	
J6188-04	1.371	0.0	0.005	
J6188-04DUP	-0.310	0.0	0.004	
J6191-02	-0.046	0.0	0.004	
J6191-04	-0.146	0.0	0.004	
J6192-04	-0.566	0.0	0.003	
J6200-02	-0.016	0.0	0.004	
J6498-07	-0.020	0.0	0.004	
J6498-08	0.019	0.0	0.004	
J6499-02	0.057	0.0	0.004	
CCV2	234.538	0.0	0.202	
CCB2	0.056	0.0	0.004	
J6500-02	-0.076	0.0	0.004	
J6501-02	-0.095	0.0	0.004	
J6518-03	-0.251	0.0	0.004	
J6520-04	-0.032	0.0	0.004	
PB115924BL	-0.082	0.0	0.004	
J6496-02	-0.166	0.0	0.004	
J6496-02DUP	-0.292	0.0	0.004	
J6500-01	-0.626	0.0	0.003	
J6503-02	-0.193	0.0	0.004	
J6516-02	69.629	0.0	0.063	
CCV3	249.039	0.0	0.215	
CCB3	-0.343	0.0	0.004	

N	28
Mean	31.524
SD	76.6920
CV%	243.28

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 21 10:31:01 2018

Fri Dec 21 12:18:03 2018

AS

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.OPPBCN	A	Reactive CN	P	-0.388	µg/l	12/21/2018 10:31:01
5.OPPBCN	A	Reactive CN	P	4.7413	µg/l	12/21/2018 10:31:02
10PPBCN	A	Reactive CN	P	9.1935	µg/l	12/21/2018 10:31:03
50PPBCN	A	Reactive CN	P	48.8982	µg/l	12/21/2018 10:31:04
100PPBCN	A	Reactive CN	P	100.9718	µg/l	12/21/2018 10:31:05
250PPBCN	A	Reactive CN	P	253.2972	µg/l	12/21/2018 10:31:06
500PPBCN	A	Reactive CN	P	498.2859	µg/l	12/21/2018 10:31:07
ICV1	S	Reactive CN	P	96.1163	µg/l	12/21/2018 11:40:57
ICB1	S	Reactive CN	P	0.042	µg/l	12/21/2018 11:40:58
CCV1	S	Reactive CN	P	235.6117	µg/l	12/21/2018 11:40:59
CCB1	S	Reactive CN	P	-0.3666	µg/l	12/21/2018 11:41:00
PB115923BL	S	Reactive CN	P	-0.1744	µg/l	12/21/2018 11:41:01
J6188-04	S	Reactive CN	P	1.3713	µg/l	12/21/2018 11:41:02
J6188-04DUP	S	Reactive CN	P	-0.3099	µg/l	12/21/2018 11:41:03
J6191-02	S	Reactive CN	P	-0.0456	µg/l	12/21/2018 11:41:04
J6191-04	S	Reactive CN	P	-0.1456	µg/l	12/21/2018 11:41:05
J6192-04	S	Reactive CN	P	-0.5662	µg/l	12/21/2018 11:41:06
J6200-02	S	Reactive CN	P	-0.0163	µg/l	12/21/2018 11:41:07
J6498-07	S	Reactive CN	P	-0.0203	µg/l	12/21/2018 11:48:30
J6498-08	S	Reactive CN	P	0.0191	µg/l	12/21/2018 11:48:31
J6499-02	S	Reactive CN	P	0.0568	µg/l	12/21/2018 11:48:32
CCV2	S	Reactive CN	P	234.5376	µg/l	12/21/2018 11:48:33
CCB2	S	Reactive CN	P	0.0563	µg/l	12/21/2018 11:48:34
J6500-02	S	Reactive CN	P	-0.0763	µg/l	12/21/2018 11:48:35
J6501-02	S	Reactive CN	P	-0.095	µg/l	12/21/2018 11:48:36
J6518-03	S	Reactive CN	P	-0.2511	µg/l	12/21/2018 11:48:37
J6520-04	S	Reactive CN	P	-0.0316	µg/l	12/21/2018 11:48:38
PB115924BL	S	Reactive CN	P	-0.082	µg/l	12/21/2018 11:48:39
J6496-02	S	Reactive CN	P	-0.1665	µg/l	12/21/2018 11:48:40
J6496-02DUP	S	Reactive CN	P	-0.2916	µg/l	12/21/2018 11:53:45
J6500-01	S	Reactive CN	P	-0.6265	µg/l	12/21/2018 11:53:46
J6503-02	S	Reactive CN	P	-0.1928	µg/l	12/21/2018 11:53:47
J6516-02	S	Reactive CN	P	69.629	µg/l	12/21/2018 11:53:48
CCV3	S	Reactive CN	P	249.0387	µg/l	12/21/2018 11:53:49
CCB3	S	Reactive CN	P	-0.3429	µg/l	12/21/2018 11:53:50

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

12/21/2018 10:38


Test Total CN

Accepted 12/21/2018 10:38

Factor 1182
 Bias 0.004

Coeff. of det. 0.999918

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.3880	0.0000	
2	5.0PPBCN	0.008	4.7413	5.0000	
3	10PPBCN	0.012	9.1935	10.0000	
4	50PPBCN	0.045	48.8982	50.0000	
5	100PPBCN	0.089	100.9718	100.0000	
6	250PPBCN	0.218	253.2972	250.0000	
7	500PPBCN	0.425	498.2859	500.0000	