

SOP ID : M3010A-Digestion-15

SDG No : N/A

Matrix : WATER

Pipette ID: ICP-A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M4462

Hood ID : #3

Block ID/Temp : 1. N/A 2. 96°C 3. N/A 4. N/A

Start Digest Date: 07/06/2020 Time : 09:25 Temp : 96 °C

End Digest Date: 07/06/2020 Time : 13:29 Temp : 96 °C

Digestion tube ID: M4578

Block thermometer ID: MET-DIG. #2

Dig Technician Signature: D.P.

Supervisor Signature: SD

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M4477
LFS-2	0.25	M4484
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
CONC: HNO3	3.00	M4580
1:1 HCL	5.00	MP56729
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Received By	Relinquished By	Location
7/6/20 13:49	SD	D.P.	Metals Lab
	Analysis Group	Preparation Group	

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
L3196-01	MW-6	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	2-1
L3196-02	MW-10	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	2
L3198-01	MW-2R	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3
L3198-02	MW-12	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	4
L3198-03	MW-1	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	5
L3198-04	MW-10	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	6
L3198-05	SW-1	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	7
L3198-05DUP	SW-1DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	8
L3198-06	L3198-05MS	<2	50	25	Colorless	Colorless	Clear	Clear	M4477,M4484	9
L3198-07	L3198-05MSD	<2	50	25	Colorless	Colorless	Clear	Clear	M4477,M4484	10
L3198-08	MW-6R	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	11
PB129951BL	PBW951	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	12
PB129951BS	LCS951	<2	50	25	Colorless	Colorless	Clear	Clear	M4477,M4484	13