


SOP ID : M3050B-Digestion-18
SDG No : N/A **Start Digest Date:** 10/23/2020 **Time :** 09:10 **Temp :** 95 °C
Matrix : SOLID **End Digest Date:** 10/23/2020 **Time :** 12:40 **Temp :** 96 °C
Pipette ID: ICP-A **Digestion tube ID:** M4577
Balance ID : M SC-2 **Block thermometer ID:** MET-DIG. #3
Filter paper ID : N/A **Dig Technician Signature:** 
pH Strip ID : N/A **Supervisor Signature:** 
Hood ID : #3
Block ID/Temp : 1. N/A 2. N/A 3. 95°C 4. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	1.00	M4568
LFS-2	1.00	M4573
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
1:1 HNO3	10.00	MP57471
CONC: HNO3	5.00	M4642
30% H2O2	3.00	M4527
CONC: HCL	10.00	M4624
PTFE Boiling Stones	-----	M4561
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Received By	Relinquished By	Location
10/23/20 12:50		D.P.	Metals Lab
	Analysis Group	Preparation Group	

Lab Sample ID	Client Sample ID	pH	Initial Weight (g)	Final Vol (ml)	Color Before	Color After	Texture	Artifact	Comment	Prep Pos
L4486-02	MS-SOIL	<2	2.24	100	Brown	Yellow	Fine	No	A	3
L4488-05	202	<2	2.31	100	Brown	Yellow	Fine	No	A	4
L4488-05MS	202MS	<2	2.29	100	Brown	Yellow	Fine	No	M4568,M4573	6
L4488-05MSD	202MSD	<2	2.22	100	Brown	Yellow	Fine	No	M4568,M4573	7
L4488-05DUP	202DUP	<2	2.32	100	Brown	Yellow	Fine	No	A	5
L4496-01	SA-01-102220	<2	2.23	100	Brown	Yellow	Fine	No	F	8
L4500-02	11935	<2	2.30	100	Brown	Yellow	Fine	No	N/A	9
L4501-01	COMP-1	<2	2.32	100	Brown	Yellow	Fine	No	N/A	10
L4501-02	COMP-2	<2	2.39	100	Brown	Yellow	Fine	No	N/A	11
L4501-03	COMP-3	<2	2.35	100	Brown	Yellow	Fine	No	N/A	12
PB132585BL	PBS585	<2	2.30	100	Colorless	Colorless	Fine	No	N/A	3-1
PB132585BS	LCS585	<2	2.29	100	Colorless	Colorless	Fine	No	M4568,M4573	2