

SOP ID : M7471B-Mercury-17

SDG No : NA

Matrix : SOIL

Pipette ID: HG-A

Balance ID : M SC-3

Filter paper ID : NA

pH Strip ID : NA

Hood ID : #1

Start Digest Date: 04/02/2021 Time : 09:25 Temp : 95 °C

End Digest Date: 04/02/2021 Time : 09:55 Temp : 94 °C

Digestion tube ID: M4685

Block thermometer ID: HG-DIG#3

Dig Technician Signature: 

Supervisor Signature: 

Block ID/Temp : 1. N/A 2. N/A 3. 95°C 4. N/A

Standardized Name	MLS USED	STD REF. # FROM LOG
ICV	30mL	MP62769
CCV	30mL	MP62771
CRA	30mL	MP62773
Blank Spike	0.48mL	MP62762
Matrix Spike	0.48mL	MP62762

Chemical Used	ML/SAMPLE USED	Lot Number
AQUA REGIA	1.5mL	MP62775
KMN04	4.5mL	MP62741
Hydroxylamine HCL	2.0mL	MP62626
PTFE Boiling Stones	-----	M4583
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

LAB SAMPLE ID	CLIENT SAMPLE ID	Wt(g)/Vol(ml)	Comment
0.0 ppb	S0	30mL	MP62763
0.05 ppb	S0.05	N/A	N/A
0.2 ppb	S0.2	30mL	MP62764
2.5 ppb	S2.5	30mL	MP62765
5.0 ppb	S5.0	30mL	MP62766
7.5 ppb	S7.5	30mL	MP62767
10.0 ppb	S10.0	30mL	MP62768
ICV	ICV	30mL	MP62769
ICB	ICB	30mL	MP62770
CCV	CCV	30mL	MP62771
CCB	CCB	30mL	MP62772
CRI	CRI	30mL	MP62773
CHK STD	CHK STD	30mL	MP62774

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Received By	Relinquished By	Location
4/2/21 10:35			metlab
	Analysis Group	Preparation Group	

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Comment	Prep Pos
M1888-01	TP-3S-(0-4)	0.58	35	NA	N/A	3-25
M1888-04	TP-3D-(4-10)	0.56	35	NA	N/A	26
M1889-01	TP-5-(0-4)	0.56	35	NA	N/A	27
M1894-01	20210155-CP-COMPOSITE	0.50	35	NA	N/A	28
M1894-02	20210156-AMENDED-CP-COMPOSITE	0.58	35	NA	N/A	29
M1897-02	TP-1	0.57	35	NA	N/A	30
M1897-02DUP	TP-1DUP	0.58	35	NA	N/A	31
M1897-02MS	TP-1MS	0.56	35	NA	MP62762	32
M1897-02MSD	TP-1MSD	0.58	35	NA	MP62762	33
M1897-04	TP-2	0.55	35	NA	N/A	34
M1900-01	CHRT-26586	0.50	35	NA	N/A	35
M1900-03	CHRT-20111	0.57	35	NA	N/A	36
M1901-01	SP-15TH-STREET	0.51	35	NA	N/A	37
M1908-01	SB-1(2-2.5)	0.57	35	NA	N/A	38
M1908-02	SB-2(2-2.5)	0.55	35	NA	N/A	39
M1908-05	SB-5(3.0-3.5)	0.52	35	NA	N/A	40
M1908-06	SB-6(3.0-3.5)	0.59	35	NA	N/A	41
M1909-01	FS-SB-10-5-10	0.54	35	NA	N/A	42
M1909-02	FS-SB-10-26-27	0.52	35	NA	N/A	43
M1909-03	FS-SB-17-12-13	0.56	35	NA	N/A	44
M1911-01	TP-6S-(0-4)	0.53	35	NA	N/A	45
M1911-04	TP-6D-(4-10)	0.59	35	NA	N/A	46
PB135507BL	PBS507	0.54	35	NA	N/A	47
PB135507BS	LCS507	0.51	35	NA	MP62762	48