

SOP ID : M3010A-Digestion-15
SDG No : N/A **Start Digest Date:** 09/03/2021 **Time :** 10:35 **Temp :** 95 °C
Matrix : WATER **End Digest Date:** 09/03/2021 **Time :** 15:06 **Temp :** 95 °C
Pipette ID: ICP-A **Digestion tube ID:** M4912
Balance ID : N/A **Block thermometer ID:** MET-DIG#2
Filter paper ID : N/A **Dig Technician Signature:** BH
pH Strip ID : M4909 **Supervisor Signature:** [Signature]
Hood ID : #3
Block ID/Temp : 1. N/A 2. 95°C 3. N/A 4. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M4812
LFS-2	0.25	M4820
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
CONC: HNO3	3.00	M4924
1:1 HCL	5.00	MP63865
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

N/A

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
09/03/21 15:45	BH / Dig Lab	SA (methods lab)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
M3606-01	BPU-1	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3
M3606-02	BPU-2	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	4
M3606-02DUP	BPU-2DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	5
M3606-03	M3606-02MS	<2	50	25	Colorless	Colorless	Clear	Clear	M4812,M4820	6
M3606-04	M3606-02MSD	<2	50	25	Colorless	Colorless	Clear	Clear	M4812,M4820	7
M3606-05	BPU-BFD	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	8
M3607-01	20211021-SITE-WATER	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	9
M3607-02	20211022-FIELD-BLANK	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	10
M3629-01	LPA-02	<2	50	25	L.Yellow	L.Yellow	Clear	Clear	N/A	11
M3648-05	20211056-SI-COMP-1-MO DIFIED ELUTRIATE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	12
M3648-06	20211056-SI-COMP-1-MO DIFIED ELUTRIATE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	13
M3648-07	20211057-SI-COMP-2-MO DIFIED ELUTRIATE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	14
M3648-08	20211057-SI-COMP-2-MO DIFIED ELUTRIATE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	15
M3648-09	20211058-SI-COMP-3-MO DIFIED ELUTRIATE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	16
M3648-10	20211058-SI-COMP-3-MO DIFIED ELUTRIATE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	17
M3648-11	20211059-SI-COMP-4-MO DIFIED ELUTRIATE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	18
M3648-12	20211059-SI-COMP-4-MO DIFIED ELUTRIATE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	19
PB138910BL	PBW910	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	2-1
PB138910BS	LCS910	<2	50	25	Colorless	Colorless	Clear	Clear	M4812,M4820	2