

Data Path : Z:\HPCHEM1\BNA B\DATA\BB030317\
 Data File : BB058688.D
 Acq On : 3 Mar 2017 16:07
 Operator : UM/SJ
 Sample : SSTD16081
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_B
 ClientSampleId :
 SSTD16081

Quant Time: Mar 15 01:57:06 2017
 Quant Method : Z:\HPCHEM1\BNA B\METHOD\SOM-EPA-BB030317-PAH.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 15 01:55:23 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	191053	20.00	ng/ul	0.01
2) Naphthalene-d8	10.83	136	667655	20.00	ng/ul	0.03
6) Acenaphthene-d10	14.81	164	459072	20.00	ng/ul	0.03
12) Phenanthrene-d10	17.66	188	750794	20.00	ng/ul	0.04
18) Chrysene-d12	22.36	240	803206	20.00	ng/ul	0.04
23) Perylene-d12	25.35	264	644753	20.00	ng/ul	0.06
System Monitoring Compounds						
3) 2,4-Dichlorophenol-d3	0.00	165	0d	0.00	ng/uL	
7) Acenaphthylene-d8	0.00	160	0d	0.00	ng/ul	
10) Fluorene-d10	0.00	176	0d	0.00	ng/ul	
15) Anthracene-d10	0.00	188	0d	0.00	ng/ul	
19) Pyrene-d10	0.00	212	0d	0.00	ng/ul	
26) Benzo(a)pyrene-d12	0.00	264	0d	0.00	ng/ul	
Target Compounds						
13) Pentachlorophenol	17.34	266	1181981	202.02	ng/uL	99
17) Fluoranthene	19.85	202	6653782	156.88	ng/ul#	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA B\DATA\BB030317\
Data File : BB058688.D
Acq On : 3 Mar 2017 16:07
Operator : UM/SJ
Sample : SSTD16081
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_B
ClientSampleId :
SSTD16081

Quant Time: Mar 15 01:57:06 2017
Quant Method : Z:\HPCHEM1\BNA B\METHOD\SOM-EPA-BB030317-PAH.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 15 01:55:23 2017
Response via : Initial Calibration

