

Data Path : Z:\HPCHEM1\BNA B\DATA\BB030317\
 Data File : BB058689.D
 Acq On : 3 Mar 2017 16:51
 Operator : UM/SJ
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_B
 ClientSampleId :
 SICV82

Quant Time: Mar 15 01:58:14 2017
 Quant Method : Z:\HPCHEM1\BNA B\METHOD\SOM-EPA-BB030317-PAH.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 15 01:57:54 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	177102	20.00	ng/ul	0.00
2) Naphthalene-d8	10.80	136	605833	20.00	ng/ul	0.00
6) Acenaphthene-d10	14.78	164	405182	20.00	ng/ul	0.00
12) Phenanthrene-d10	17.62	188	657329	20.00	ng/ul	0.00
18) Chrysene-d12	22.32	240	549630	20.00	ng/ul	0.00
23) Perylene-d12	25.28	264	531954	20.00	ng/ul	-0.01
System Monitoring Compounds						
3) 2,4-Dichlorophenol-d3	10.42	165	234624	20.24	ng/uL	0.00
7) Acenaphthylene-d8	14.46	160	691203	19.77	ng/ul	0.00
10) Fluorene-d10	15.81	176	470730	19.62	ng/ul	0.00
15) Anthracene-d10	17.73	188	637256	20.61	ng/ul	0.00
19) Pyrene-d10	20.19	212	575245	21.32	ng/ul	0.00
26) Benzo(a)pyrene-d12	25.09	264	486270	20.72	ng/ul	-0.01
Target Compounds				Qvalue		
4) Naphthalene	10.85	128	582992	20.05	ng/ul	99
5) 2-Methylnaphthalene	12.52	142	439008	20.08	ng/ul	100
8) Acenaphthylene	14.49	152	675128	20.13	ng/ul	99
9) Acenaphthene	14.85	153	443893	19.71	ng/ul	100
11) Fluorene	15.87	166	573540	19.61	ng/ul	100
13) Pentachlorophenol	17.29	266	101843	19.05	ng/uL	99
14) Phenanthrene	17.67	178	768271	20.64	ng/ul	100
16) Anthracene	17.77	178	791071	20.75	ng/ul	99
17) Fluoranthene	19.82	202	754706	20.32	ng/ul	99
20) Pyrene	20.22	202	776653	21.39	ng/ul	99
21) Benzo(a)anthracene	22.30	228	651790	21.46	ng/ul	99
22) Chrysene	22.37	228	594400	21.24	ng/ul	99
24) Benzo(b)fluoranthene	24.38	252	628730	20.23	ng/ul	98
25) Benzo(k)fluoranthene	24.45	252	589631	20.94	ng/ul	100
27) Benzo(a)pyrene	25.16	252	589903	20.21	ng/ul	99
28) Indeno(1,2,3-cd)pyrene	28.37	276	675806	20.25	ng/ul	98
29) Dibenzo(a,h)anthracene	28.40	278	587783	20.72	ng/ul	99
30) Benzo(g,h,i)perylene	29.30	276	568287	20.63	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA B\DATA\BB030317\
Data File : BB058689.D
Acq On : 3 Mar 2017 16:51
Operator : UM/SJ
Sample : SSTDICV020
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_B
ClientSampleId :
SICV82

Quant Time: Mar 15 01:58:14 2017
Quant Method : Z:\HPCHEM1\BNA B\METHOD\SOM-EPA-BB030317-PAH.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 15 01:57:54 2017
Response via : Initial Calibration

