

Data Path : Z:\HPCHEM1\BNA B\DATA\BB030917\
 Data File : BB058723.D
 Acq On : 9 Mar 2017 21:02
 Operator : UM/SJ
 Sample : I2107-17
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_B
 ClientSampleId :
 C0AL3

Quant Time: Mar 15 15:50:43 2017
 Quant Method : Z:\HPCHEM1\BNA B\METHOD\SOM-EPA-BB030917-PAH.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 15 14:21:30 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	4.87	152	303745	20.00	ng/ul	-0.02
2) Naphthalene-d8	6.15	136	1077003	20.00	ng/ul	-0.02
6) Acenaphthene-d10	8.01	164	751851	20.00	ng/ul	-0.03
12) Phenanthrene-d10	9.68	188	1443464	20.00	ng/ul	0.04
18) Chrysene-d12	12.73	240	1484092	20.00	ng/ul	-0.04
23) Perylene-d12	15.63	264	1095467	20.00	ng/ul	-0.06
System Monitoring Compounds						
3) 2,4-Dichlorophenol-d3	5.95	165	345735	16.85	ng/uL	-0.03
7) Acenaphthylene-d8	7.85	160	1325934	22.63	ng/ul	-0.03
10) Fluorene-d10	8.56	176	981873	20.63	ng/ul	-0.02
15) Anthracene-d10	9.68	188	1443464	22.48	ng/ul	-0.02
19) Pyrene-d10	11.16	212	1546448	22.91	ng/ul	-0.03
26) Benzo(a)pyrene-d12	15.43	264	1015818	20.73	ng/ul	-0.06
Target Compounds				Qvalue		
20) Pyrene	11.19	202	160183	1.94	ng/ul#	90
24) Benzo(b)fluoranthene	14.70	252	75683	1.13	ng/ul	95
25) Benzo(k)fluoranthene	14.70	252	75683	1.29	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA B\DATA\BB030917\
Data File : BB058723.D
Acq On : 9 Mar 2017 21:02
Operator : UM/SJ
Sample : I2107-17
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_B
ClientSampleId :
C0AL3

Quant Time: Mar 15 15:50:43 2017
Quant Method : Z:\HPCHEM1\BNA B\METHOD\SOM-EPA-BB030917-PAH.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 15 14:21:30 2017
Response via : Initial Calibration

