

Data Path : Z:\HPCHEM1\BNA_B\DATA\BB101916\
 Data File : BB058652.D
 Acq On : 19 Oct 2016 14:41
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_B
 LabSampleId :
 SSTD02096

Quant Time: Oct 20 12:09:32 2016
 Quant Method : Z:\HPCHEM1\BNA_B\METHOD\SOM02.2-EPA-BB101916-MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 19 14:48:40 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2 S	1,4-Dioxane-d8	0.460	0.441	4.1	90	0.00
3 S	Phenol-d5	1.531	1.432	6.5	90	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.078	0.989	8.3	81	0.00
5 S	2-Chlorophenol-d4	1.519	1.498	1.4	94	0.00
6 S	4-Methylphenol-d8	1.233	1.171	5.0	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	0.02
8 S	Nitrobenzene-d5	0.200	0.203	-1.5	92	0.00
9 S	2-Nitrophenol-d4	0.217	0.231	-6.5	96	0.00
10 S	2,4-Dichlorophenol-d3	0.299	0.321	-7.4	101	0.00
11 S	4-Chloroaniline-d4	0.383	0.415	-8.4	96	0.00
12	1-Methylnaphthalene	0.633	0.681	-7.6	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.02
14 S	Dimethylphthalate-d6	1.324	1.367	-3.2	98	0.00
15 S	Acenaphthylene-d8	1.557	1.640	-5.3	97	0.00
16 S	4-Nitrophenol-d4	0.299	0.276	7.7	95	0.00
17 S	Fluorene-d10	1.079	1.155	-7.0	97	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
19 S	4,6-Dinitro-2-methylphenol-	0.210	0.191	9.0	100	0.00
20 S	Anthracene-d10	0.909	0.989	-8.8	100	0.12
21	1,2,3,4-Tetrachlorobenzene	0.353	0.368	-4.2	99	0.00
22	Pentachlorobenzene	0.345	0.361	-4.6	99	0.00
23 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
24 S	Pyrene-d10	0.891	0.992	-11.3	100	0.00
25 I	Perylene-d12	1.000	1.000	0.0	100	0.02
26 S	Benzo(a)pyrene-d12	0.901	0.933	-3.6	104	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0