

Data Path : Z:\HPCHEM1\BNA_B\DATA\BB101916\
 Data File : BB058661.D
 Acq On : 19 Oct 2016 20:52
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_B
 LabSampleId :
 SSTD02097

Quant Time: Oct 20 16:21:33 2016
 Quant Method : Z:\HPCHEM1\BNA_B\METHOD\SOM02.2-EPA-BB101916-MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 20 04:20:07 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2 S	1,4-Dioxane-d8	0.460	0.482	-4.8	100	0.00
3 S	Phenol-d5	1.531	1.612	-5.3	103	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.078	1.235	-14.6	103	0.00
5 S	2-Chlorophenol-d4	1.519	1.584	-4.3	101	0.00
6 S	4-Methylphenol-d8	1.233	1.281	-3.9	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.200	0.208	-4.0	99	0.00
9 S	2-Nitrophenol-d4	0.217	0.228	-5.1	99	0.00
10 S	2,4-Dichlorophenol-d3	0.299	0.308	-3.0	102	0.00
11 S	4-Chloroaniline-d4	0.383	0.420	-9.7	102	0.00
12	1-Methylnaphthalene	0.633	0.680	-7.4	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.02
14 S	Dimethylphthalate-d6	1.324	1.379	-4.2	99	0.00
15 S	Acenaphthylene-d8	1.557	1.687	-8.3	99	0.00
16 S	4-Nitrophenol-d4	0.299	0.292	2.3	100	0.00
17 S	Fluorene-d10	1.079	1.184	-9.7	99	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
19 S	4,6-Dinitro-2-methylphenol-	0.210	0.186	11.4	97	0.00
20 S	Anthracene-d10	0.909	0.973	-7.0	98	0.10
21	1,2,3,4-Tetrachlorobenzene	0.353	0.353	0.0	95	0.00
22	Pentachlorobenzene	0.345	0.355	-2.9	97	0.00
23 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
24 S	Pyrene-d10	0.891	0.942	-5.7	96	0.00
25 I	Perylene-d12	1.000	1.000	0.0	91	0.00
26 S	Benzo(a)pyrene-d12	0.901	0.916	-1.7	93	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0