

Data Path : Z:\HPCHEM1\BNA_B\Data\BB102016\
 Data File : BB058672.D
 Acq On : 20 Oct 2016 9:21
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_B
 LabSampleId :
 SSTD02067

Quant Time: Oct 20 16:26:25 2016
 Quant Method : Z:\HPCHEM1\BNA_B\METHOD\SOM02.2-EPA-BB101916-MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 20 16:25:04 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2 S	1,4-Dioxane-d8	0.460	0.524	-13.9	104	0.00
3 S	Phenol-d5	1.531	1.665	-8.8	102	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.078	1.300	-20.6#	105	0.00
5 S	2-Chlorophenol-d4	1.519	1.587	-4.5	97	0.00
6 S	4-Methylphenol-d8	1.233	1.315	-6.7	100	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	93	0.02
8 S	Nitrobenzene-d5	0.200	0.209	-4.5	97	0.00
9 S	2-Nitrophenol-d4	0.217	0.217	0.0	93	0.00
10 S	2,4-Dichlorophenol-d3	0.299	0.293	2.0	95	0.00
11 S	4-Chloroaniline-d4	0.383	0.422	-10.2	101	0.00
12	1-Methylnaphthalene	0.633	0.685	-8.2	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.02
14 S	Dimethylphthalate-d6	1.324	1.318	0.5	94	0.00
15 S	Acenaphthylene-d8	1.557	1.663	-6.8	97	0.00
16 S	4-Nitrophenol-d4	0.299	0.280	6.4	95	0.00
17 S	Fluorene-d10	1.079	1.148	-6.4	95	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
19 S	4,6-Dinitro-2-methylphenol-	0.210	0.172	18.1	89	0.00
20 S	Anthracene-d10	0.909	0.920	-1.2	92	0.10
21	1,2,3,4-Tetrachlorobenzene	0.353	0.347	1.7	92	0.00
22	Pentachlorobenzene	0.345	0.347	-0.6	94	0.00
23 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
24 S	Pyrene-d10	0.891	0.990	-11.1	95	0.00
25 I	Perylene-d12	1.000	1.000	0.0	84	0.00
26 S	Benzo(a)pyrene-d12	0.901	0.917	-1.8	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0