

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE010219\
 Data File : BE098600.D
 Acq On : 2 Jan 2019 9:44
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 02 10:34:04 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	63	0.00
2	1,4-Dioxane	0.692	0.784	-13.3	72	0.00
3	n-Nitrosodimethylamine	0.623	0.654	-5.0	71	0.00
4 S	2-Fluorophenol	0.936	1.107	-18.3	76	0.01
5 S	Phenol-d6	1.326	1.736	-30.9#	82	0.01
6	bis(2-Chloroethyl)ether	1.260	1.224	2.9	63	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
8 S	Nitrobenzene-d5	0.364	0.378	-3.8	74	0.00
9	Naphthalene	4.025	4.222	-4.9	71	0.00
10	Hexachlorobutadiene	0.256	0.292	-14.1	77	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.701	-7.8	72	0.00
12	2-Methylnaphthalene	0.654	0.687	-5.0	71	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	66	0.00
14 S	2,4,6-Tribromophenol	0.147	0.195	-32.7#	92	0.01
15 S	2-Fluorobiphenyl	1.687	1.658	1.7	71	0.02
16	Acenaphthylene	1.874	1.787	4.6	65	0.02
17	Acenaphthene	1.126	1.167	-3.6	69	0.02
18	Fluorene	1.506	1.528	-1.5	67	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	58	0.00
20	4-Bromophenyl-phenylether	0.215	0.226	-5.1	63	0.01
21	Hexachlorobenzene	0.231	0.259	-12.1	66	0.01
22	Pentachlorophenol	0.046	0.115	-150.0#	183#	0.00
23	Phenanthrene	0.972	1.009	-3.8	61	0.01
24	Anthracene	0.866	0.865	0.1	59	0.01
25 SURR	Fluoranthene-d10	5.131	5.329	-3.9	58	0.00
26	Fluoranthene	1.286	1.291	-0.4	56	0.01
27 I	Chrysene-d12	1.000	1.000	0.0	63	0.01
28	Pvrene	1.255	1.188	5.3	58	0.01
29 S	Terphenyl-d14	0.972	0.951	2.2	62	0.00
30	Benzo(a)anthracene	1.139	1.208	-6.1	69	0.00
31	Chrysene	1.193	1.173	1.7	63	0.01
32	Bis(2-ethylhexyl)phthalate	2.313	2.839	-22.7	78	0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.226	-3.3	69	-0.06
34 I	Perylene-d12	1.000	1.000	0.0	66	-0.03
35	Benzo(b)fluoranthene	1.094	1.166	-6.6	73	-0.01
36	Benzo(k)fluoranthene	1.274	1.268	0.5	69	-0.02
37 C	Benzo(a)pyrene	1.129	1.151	-1.9	69	-0.03
38	Dibenzo(a,h)anthracene	1.063	1.029	3.2	66	-0.08
39	Benzo(a,h,i)perylene	1.071	1.086	-1.4	69	-0.08

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE010219\
Data File : BE098600.D
Acq On : 2 Jan 2019 9:44
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 02 10:34:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			