

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010419\
 Data File : BE098614.D
 Acq On : 4 Jan 2019 12:23
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 04 21:12:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 02 10:30:27 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	64	0.00
2	1,4-Dioxane	0.692	0.699	-1.0	66	0.00
3	n-Nitrosodimethylamine	0.623	0.638	-2.4	71	0.01
4 S	2-Fluorophenol	0.936	1.243	-32.8#	88	0.01
5 S	Phenol-d6	1.326	1.853	-39.7#	89	0.01
6	bis(2-Chloroethyl)ether	1.260	1.266	-0.5	67	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
8 S	Nitrobenzene-d5	0.364	0.387	-6.3	76	0.00
9	Naphthalene	4.025	4.281	-6.4	72	0.00
10	Hexachlorobutadiene	0.256	0.297	-16.0	79	0.00
11 SURR	2-Methylnaphthalene-d10	0.650	0.723	-11.2	74	0.00
12	2-Methylnaphthalene	0.654	0.717	-9.6	74	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	67	0.00
14 S	2,4,6-Tribromophenol	0.147	0.186	-26.5#	90	0.01
15 S	2-Fluorobiphenyl	1.687	1.659	1.7	72	0.02
16	Acenaphthylene	1.874	1.860	0.7	69	0.02
17	Acenaphthene	1.126	1.174	-4.3	71	0.02
18	Fluorene	1.506	1.514	-0.5	68	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	56	0.00
20	4-Bromophenyl-phenylether	0.215	0.231	-7.4	63	0.01
21	Hexachlorobenzene	0.231	0.266	-15.2	66	0.01
22	Pentachlorophenol	0.046	0.103	-123.9#	160#	0.01
23	Phenanthrene	0.972	1.025	-5.5	60	0.01
24	Anthracene	0.866	0.891	-2.9	59	0.01
25 SURR	Fluoranthene-d10	5.131	5.218	-1.7	55	0.01
26	Fluoranthene	1.286	1.276	0.8	54	0.01
27 I	Chrysene-d12	1.000	1.000	0.0	61	0.01
28	Pvrene	1.255	1.197	4.6	56	0.01
29 S	Terphenyl-d14	0.972	0.937	3.6	59	0.01
30	Benzo(a)anthracene	1.139	1.211	-6.3	66	0.00
31	Chrysene	1.193	1.211	-1.5	62	0.01
32	Bis(2-ethylhexyl)phthalate	2.313	3.049	-31.8#	80	0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.286	-8.3	69	-0.06
34 I	Perylene-d12	1.000	1.000	0.0	66	-0.03
35	Benzo(b)fluoranthene	1.094	1.148	-4.9	71	-0.01
36	Benzo(k)fluoranthene	1.274	1.239	2.7	67	-0.01
37 C	Benzo(a)pyrene	1.129	1.127	0.2	67	-0.03
38	Dibenzo(a,h)anthracene	1.063	1.050	1.2	67	-0.07
39	Benzo(a,h,i)perylene	1.071	1.114	-4.0	71	-0.08

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010419\
Data File : BE098614.D
Acq On : 4 Jan 2019 12:23
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 04 21:12:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 02 10:30:27 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			