

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010920\
 Data File : BE101010.D
 Acq On : 9 Jan 2020 12:54
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 09 18:11:44 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA E\Methods\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	1.020	0.903	11.5	96	0.00
3	n-Nitrosodimethylamine	0.556	0.502	9.7	105	0.00
4 S	2-Fluorophenol	0.900	0.838	6.9	95	0.00
5 S	Phenol-d6	1.139	0.989	13.2	98	0.00
6	bis(2-Chloroethyl)ether	1.357	1.146	15.5	93	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.01
8 S	Nitrobenzene-d5	0.286	0.239	16.4	96	0.00
9	Naphthalene	4.441	4.569	-2.9	103	0.00
10	Hexachlorobutadiene	0.188	0.201	-6.9	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.763	0.775	-1.6	104	0.00
12	2-Methylnaphthalene	0.672	0.665	1.0	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.113	0.099	12.4	100	0.00
15 S	2-Fluorobiphenyl	1.519	1.569	-3.3	101	0.00
16	Acenaphthylene	1.629	1.464	10.1	99	0.00
17	Acenaphthene	1.166	1.162	0.3	102	0.00
18	Fluorene	1.474	1.449	1.7	100	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.195	0.188	3.6	99	0.00
21	Hexachlorobenzene	0.214	0.222	-3.7	100	0.00
22	Pentachlorophenol	0.053	0.033	37.7#	79	0.00
23	Phenanthrene	1.092	1.075	1.6	98	0.00
24	Anthracene	1.031	0.980	4.9	98	0.00
25 SURR	Fluoranthene-d10	5.555	5.307	4.5	97	0.00
26	Fluoranthene	1.368	1.299	5.0	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
28	Pvrene	1.489	1.551	-4.2	96	0.00
29 S	Terphenyl-d14	0.982	0.996	-1.4	95	0.00
30	Benzo(a)anthracene	1.141	0.922	19.2	80	0.00
31	Chrysene	1.696	1.961	-15.6	107	0.00
32	Bis(2-ethylhexyl)phthalate	1.902	1.561	17.9	89	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	1.431	-6.8	104	0.00
34 I	Perylene-d12	1.000	1.000	0.0	107	0.00
35	Benzo(b)fluoranthene	1.122	0.930	17.1	95	0.00
36	Benzo(k)fluoranthene	1.626	1.385	14.8	91	0.00
37 C	Benzo(a)pyrene	1.085	1.010	6.9	103	0.00
38	Dibenzo(a,h)anthracene	0.998	0.921	7.7	107	0.00
39	Benzo(a,h,i)perylene	1.227	1.113	9.3	94	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE010920\
Data File : BE101010.D
Acq On : 9 Jan 2020 12:54
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 09 18:11:44 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA E\Methods\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			